

LB109573


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47640



A= 0.0000e+000

B= 2.0992e-004

C= 1.4859e-002

Rho= 0.9999886

Accept=Accepted

slope
y=mx+b

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.002	-0.002	-79	0.000	-79				
0.05	0.050					223				
0.20	0.200	0.191	-0.009	838	0.0 %	838				
2.50	2.500	2.520	0.020	11935	0.0 %	11935				
5.00	5.000	5.008	0.008	23785	0.0 %	23785				
7.50	7.500	7.467	-0.033	35502	0.0 %	35502				
10.0	10.000	10.015	0.015	47640	0.0 %	47640				

4/6/15
-5
1
000
0

LB109573
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-79	-		0 ISM02.4	PPB	6/12/2020 11:19	Std
	0.2 S0.2	838	-		0.2 ISM02.4	PPB	6/12/2020 11:22	Std
	2.5 S2.5	11935	-		2.5 ISM02.4	PPB	6/12/2020 11:24	Std
	5 S5.0	23785	-		5 ISM02.4	PPB	6/12/2020 11:26	Std
	7.5 S7.5	35502	-		7.5 ISM02.4	PPB	6/12/2020 11:28	Std
	10 S10.0	47640	-		10 ISM02.4	PPB	6/12/2020 11:31	Std
ICV81	ICV81	18592	3.9177 -		ISM02.4	PPB	6/12/2020 11:33	SMPL
ICB81	ICB81	-183	-0.0236 -		ISM02.4	PPB	6/12/2020 11:36	SMPL
CCV005	CCV005	23275	4.9007 -		ISM02.4	PPB	6/12/2020 11:38	SMPL
CCB005	CCB005	-174	-0.0217 -		ISM02.4	PPB	6/12/2020 11:40	SMPL
PB129512BL	PBW512	-25	0.0096 -		ISM02.4	PPB	6/12/2020 11:42	SMPL
L2907-01	MCOAE3	-100	-0.0061 -		ISM02.4	PPB	6/12/2020 11:45	SMPL
L2907-02	MCOAE3D	-110	-0.0082 -		ISM02.4	PPB	6/12/2020 11:47	SMPL
L2907-03	MCOAE3S	4758	1.0137 -		ISM02.4	PPB	6/12/2020 11:49	SMPL
PB129513BL	PBW513	-26	0.0094 -		ISM02.4	PPB	6/12/2020 11:51	SMPL
L2909-01	MBFSX7	-23	0.01 -		ISM02.4	PPB	6/12/2020 11:54	SMPL
L2909-02	MBFSX8	-137	-0.0139 -		ISM02.4	PPB	6/12/2020 11:56	SMPL
L2909-03	MBFSX9	-106	-0.0074 -		ISM02.4	PPB	6/12/2020 11:58	SMPL
L2909-04	MBFSZ2	-77	-0.0013 -		ISM02.4	PPB	6/12/2020 12:00	SMPL
L2909-05	MBFT03	-190	-0.025 -		ISM02.4	PPB	6/12/2020 12:03	SMPL
L2909-06	MBFSX3	-178	-0.0225 -		ISM02.4	PPB	6/12/2020 12:05	SMPL
L2909-07	MBFSX3D	-174	-0.0217 -		ISM02.4	PPB	6/12/2020 12:07	SMPL
L2909-08	MBFSX3S	4825	1.0277 -		ISM02.4	PPB	6/12/2020 12:10	SMPL
L2909-09	MBFSX6	277	0.073 -		ISM02.4	PPB	6/12/2020 12:12	SMPL
L2909-10	MBFT04	-189	-0.0248 -		ISM02.4	PPB	6/12/2020 12:14	SMPL
L2909-11	MBFSX5	-159	-0.0185 -		ISM02.4	PPB	6/12/2020 12:16	SMPL
L2909-12	MBFSY0	-165	-0.0198 -		ISM02.4	PPB	6/12/2020 12:19	SMPL
L2909-13	MBFSY1	-179	-0.0227 -		ISM02.4	PPB	6/12/2020 12:21	SMPL
L2909-14	MBFT06	-181	-0.0231 -		ISM02.4	PPB	6/12/2020 12:23	SMPL
PB129514BL	PBW514	-111	-0.0084 -		ISM02.4	PPB	6/12/2020 12:25	SMPL
CCV006	CCV006	22832	4.8078 -		ISM02.4	PPB	6/12/2020 12:28	SMPL
CCB006	CCB006	-157	-0.0181 -		ISM02.4	PPB	6/12/2020 12:30	SMPL
L2953-01	MBFTZ0	-168	-0.0204 -		ISM02.4	PPB	6/12/2020 12:32	SMPL
L2953-02	MBFTZ1	-155	-0.0177 -		ISM02.4	PPB	6/12/2020 12:34	SMPL
L2953-03	MBFTZ2	-163	-0.0194 -		ISM02.4	PPB	6/12/2020 12:37	SMPL
L2953-04	MBFTZ2D	-155	-0.0177 -		ISM02.4	PPB	6/12/2020 12:39	SMPL
L2953-05	MBFTZ2S	4560	0.9721 -		ISM02.4	PPB	6/12/2020 12:41	SMPL
L2953-06	MBFTZ3	-186	-0.0242 -		ISM02.4	PPB	6/12/2020 12:43	SMPL
L2953-07	MBFTZ4	-158	-0.0183 -		ISM02.4	PPB	6/12/2020 12:46	SMPL
L2953-08	MBFTZ5	-157	-0.0181 -		ISM02.4	PPB	6/12/2020 12:48	SMPL
L2953-09	MBFTZ6	-153	-0.0173 -		ISM02.4	PPB	6/12/2020 12:50	SMPL
L2953-10	MBFTZ7	-159	-0.0185 -		ISM02.4	PPB	6/12/2020 12:53	SMPL
L2953-11	MBFTZ8	-161	-0.0189 -		ISM02.4	PPB	6/12/2020 12:55	SMPL
L2953-12	MBFW00	-169	-0.0206 -		ISM02.4	PPB	6/12/2020 12:57	SMPL
L2953-13	MBFW01	-165	-0.0198 -		ISM02.4	PPB	6/12/2020 12:59	SMPL
L2953-14	MBFW02	-159	-0.0185 -		ISM02.4	PPB	6/12/2020 13:02	SMPL

LB109573

INSTRUMENT ID : CV1

L2953-15	MBFW04	-137	-0.0139 -	ISM02.4	PPB	6/12/2020 13:04 SMPL
L2953-16	MBFW05	-162	-0.0191 -	ISM02.4	PPB	6/12/2020 13:06 SMPL
L2953-17	MBFW06	-160	-0.0187 -	ISM02.4	PPB	6/12/2020 13:08 SMPL
L2953-18	MBFW07	-161	-0.0189 -	ISM02.4	PPB	6/12/2020 13:11 SMPL
L2953-19	MBFW08	-162	-0.0191 -	ISM02.4	PPB	6/12/2020 13:13 SMPL
L2953-20	MBFW09	-164	-0.0196 -	ISM02.4	PPB	6/12/2020 13:15 SMPL
CCV007	CCV007	22494	4.7368 -	ISM02.4	PPB	6/12/2020 13:17 SMPL
CCB007	CCB007	-176	-0.0221 -	ISM02.4	PPB	6/12/2020 13:20 SMPL
L2953-21	MBFW10	-148	-0.0162 -	ISM02.4	PPB	6/12/2020 13:22 SMPL
CCV008	CCV008	22223	4.6799 -	ISM02.4	PPB	6/12/2020 13:24 SMPL
CCB008	CCB008	-187	-0.0244 -	ISM02.4	PPB	6/12/2020 13:27 SMPL