



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109580

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/12/2020	11:30
2	CAL2	1	Water	NA	NA	20.2	7.00	06/12/2020	11:31
3	CAL3	1	Water	NA	NA	20.2	10.03	06/12/2020	11:32
4	ICV	1	Water	NA	NA	20.2	7.00	06/12/2020	11:35
5	CCV1	1	Water	NA	NA	20.1	2.02	06/12/2020	11:36
6	L3002-01	1	Water	NA	NA	21.7	5.07	06/12/2020	11:39
7	L3002-01DUP	1	Water	NA	NA	21.8	5.08	06/12/2020	11:40
8	CCV2	1	Water	NA	NA	20.2	12.02	06/12/2020	11:42

VB209580

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3002 WorkList ID : 139591 Department : Wet-Chemistry Date : 06-12-2020 11:22:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/11/2020	Water	L3002-01	pH	Cool 4 deg C	STER14	P22	JS-SYSTEM	06/11/2020	SM 4500-H B

Date/Time 06-12-2020 12:13 0
Received by: JR
Relinquished by: JR