


LB109584

245.1

INSTRUMENT ID: CXT



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	314	0.000	314				
0.05	0.050	0.054	0.004	519	0.0 %	519				
0.2	0.200	0.209	0.009	1146	0.0 %	1146				
2.5	2.500	2.550	0.050	10588	0.0 %	10588				
5.0	5.000	5.007	0.007	20498	0.0 %	20498				
7.5	7.500	7.297	-0.203	29737	0.0 %	29737				
10.0	10.000	10.130	0.130	41164	0.0 %	41164				

LB109584 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Jun 2020 13:56:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	314	-	0.0000	245.1	PPB	12 Jun 2020 13:59:05	S	Std
0.05 - 1	50.05	519	-	0.0500	245.1	PPB	12 Jun 2020 14:01:22	S	Std
0.2 - 1	50.2	1146	-	0.2000	245.1	PPB	12 Jun 2020 14:03:40	S	Std
2.5 - 1	52.5	10588	-	2.5000	245.1	PPB	12 Jun 2020 14:05:57	S	Std
5.0 - 1	55	20498	-	5.0000	245.1	PPB	12 Jun 2020 14:08:13	S	Std
7.5 - 1	57.5	29737	-	7.5000	245.1	PPB	12 Jun 2020 14:10:29	S	Std
10.0 - 1	510	41164	-	10.0000	245.1	PPB	12 Jun 2020 14:28:23	S	Std
ICV81 - 1	ICV81	15967	3.8835	-245.1		PPB	12 Jun 2020 14:33:40	U	SMPL
ICB81 - 1	ICB81	102	-0.0495	-245.1		PPB	12 Jun 2020 14:35:55	U	SMPL
CCV09 - 1	CCV09	20339	4.9673	-245.1		PPB	12 Jun 2020 14:38:10	U	SMPL
CCB09 - 1	CCB09	98	-0.0505	-245.1		PPB	12 Jun 2020 14:40:25	U	SMPL
CRA - 1	CRA	491	0.0469	-245.1		PPB	12 Jun 2020 15:00:46	U	SMPL
HighStd - 1	HighStd	41254	10.1523	-245.1		PPB	12 Jun 2020 15:03:01	U	SMPL
ChkStd - 1	ChkStd	27361	6.7081	-245.1		PPB	12 Jun 2020 15:05:16	U	SMPL
PB129522BL - 1	PBW	91	-0.0523	-245.1		PPB	12 Jun 2020 15:10:10	U	SMPL
PB129522BS - 1	LCSW	14666	3.5610	-245.1		PPB	12 Jun 2020 15:15:11	U	SMPL
L2883-01 - 1	COMP	-94	-0.0981	-245.1		PPB	12 Jun 2020 15:17:27	U	SMPL
L2883-01DUP - 1	COMPDUP	-43	-0.0855	-245.1		PPB	12 Jun 2020 15:19:43	U	SMPL
L2883-01MS - 1	COMPMS	14408	3.4970	-245.1		PPB	12 Jun 2020 15:26:55	U	SMPL
L2883-01MSD - 1	COMPMSD	14200	3.4454	-245.1		PPB	12 Jun 2020 15:29:12	U	SMPL
CCV10 - 1	CCV10	20491	5.0050	-245.1		PPB	12 Jun 2020 15:31:27	U	SMPL
CCB10 - 1	CCB10	90	-0.0525	-245.1		PPB	12 Jun 2020 15:33:42	U	SMPL
L2883-01LX5 - 1		138	-0.0406	-245.1		PPB	12 Jun 2020 15:35:58	U	SMPL
L2883-01A - 1		15128	3.6755	-245.1		PPB	12 Jun 2020 15:38:13	U	SMPL
CCV11 - 1	CCV11	19763	4.8245	-245.1		PPB	12 Jun 2020 15:40:28	U	SMPL
CCB11 - 1	CCB11	181	-0.0300	-245.1		PPB	12 Jun 2020 15:42:43	U	SMPL