



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109588

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/15/2020	08:10
2	CAL2	1	Water	NA	NA	20.1	7.00	06/15/2020	08:11
3	CAL3	1	Water	NA	NA	20.1	10.03	06/15/2020	08:12
4	ICV	1	Water	NA	NA	20.1	7.02	06/15/2020	08:13
5	CCV1	1	Water	NA	NA	20.1	2.01	06/15/2020	08:15
6	L2999-01	1	Water	NA	NA	21.1	5.02	06/15/2020	08:16
7	L2999-01DUP	1	Water	NA	NA	21.2	5.03	06/15/2020	08:17
8	CCV2	1	Water	NA	NA	20.1	12.02	06/15/2020	08:20

WORKLIST(Hardcopy Internal Chain)

18109588

WorkList Name : PH L2999 W WorkList ID : 139601 Department : Wet-Chemistry Date : 06-15-2020 07:56:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/12/2020	Water	L2999-01~C	pH	Cool 4 deg C	PSEG01	P22	31030	06/12/2020	9040C

Date/Time 06-15-2020 8:00
 Received by: Ja
 Relinquished by: Ja

Date/Time 06-15-2020 9:00
 Received by: Ja
 Relinquished by: Ja