



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109589

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/15/2020	08:10
2	CAL2	1	Water	NA	NA	20.1	7.00	06/15/2020	08:11
3	CAL3	1	Water	NA	NA	20.1	10.03	06/15/2020	08:12
4	ICV	1	Water	NA	NA	20.1	7.02	06/15/2020	08:13
5	CCV1	1	Water	NA	NA	20.1	2.01	06/15/2020	08:15
6	L2999-03	1	Solid	20.03	20	21.0	5.89	06/15/2020	08:24
7	L2999-03DUP	1	Solid	20.04	20	21.1	5.90	06/15/2020	08:25
8	CCV2	1	Water	NA	NA	20.2	12.02	06/15/2020	08:26

WORKLIST(Hardcopy Internal Chain)

RB109589

WorkList Name : PH L2999 S WorkList ID : 139602 Department : Wet-Chemistry Date : 06-15-2020 07:57:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/12/2020	Solid	L2999-03 - 1	pH	Cool 4 deg C	PSEG01	P22	TRE-1013	06/12/2020	9045D

Date/Time 06.15.2020 8:00
 Received by: Jn
 Relinquished by: J1

Date/Time 06.15.2020 9:00
 Received by: Jn
 Relinquished by: J2