

LB109600


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43973



A= 0.0000e+000

B= 2.3048e-004

C= -4.0998e-002

Rho= 0.9998043

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	272	0.000	272				
0.2	0.200	0.253	0.053	1274	0.0 %	1274				
2.5	2.500	2.510	0.010	11070	0.0 %	11070				
5.0	5.000	4.867	-0.133	21295	0.0 %	21295				
7.5	7.500	7.454	-0.046	32518	0.0 %	32518				
10.0	10.000	10.094	0.094	43973	0.0 %	43973				

LB109600 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jun 2020 14:19:23

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SC	272	-	0.00007470A	PPB	15 Jun 2020 14:22:03	S	Std
0.2 - 1	SC.2	1274	-	0.20007470A	PPB	15 Jun 2020 14:24:11	S	Std
2.5 - 1	SC.2	11070	-	2.50007470A	PPB	15 Jun 2020 14:26:19	S	Std
5.0 - 1	SC.5	21295	-	5.00007470A	PPB	15 Jun 2020 14:28:27	S	Std
7.5 - 1	SC.5	32518	-	7.50007470A	PPB	15 Jun 2020 14:30:34	S	Std
10.0 - 1	SC.5	43973	-	10.00007470A	PPB	15 Jun 2020 14:35:18	S	Std
ICV84 - 1	ICV84	17235	3.9314	-7470A	PPB	15 Jun 2020 14:40:16	U	SMPL
ICB84 - 1	ICB84	135	-0.0099	-7470A	PPB	15 Jun 2020 14:42:21	U	SMPL
CCV19 - 1	CCV19	21117	4.8262	-7470A	PPB	15 Jun 2020 14:44:27	U	SMPL
CCB19 - 1	CCB19	140	-0.0087	-7470A	PPB	15 Jun 2020 14:46:33	U	SMPL
CRA - 1	CRA	1202	0.2360	-7470A	PPB	15 Jun 2020 14:48:39	U	SMPL
HighStd - 1	HighStd	42817	9.8277	-7470A	PPB	15 Jun 2020 14:50:45	U	SMPL
ChkStd - 1	ChkStd	29489	6.7558	-7470A	PPB	15 Jun 2020 15:02:06	U	SMPL
PB129545BL - 1	PBW	104	-0.0170	-7470A	PPB	15 Jun 2020 15:04:11	U	SMPL
PB129545BS - 1	LCSW	16426	3.7449	-7470A	PPB	15 Jun 2020 15:06:19	U	SMPL
L2808-05 - 1	HR-06-061120	461	0.0653	-7470A	PPB	15 Jun 2020 15:08:26	U	SMPL
L2808-05DUP - 1	HR-06-061120DUP	513	0.0772	-7470A	PPB	15 Jun 2020 15:13:07	U	SMPL
L2808-05MS - 1	HR-06-061120MS	16977	3.8719	-7470A	PPB	15 Jun 2020 15:15:15	U	SMPL
L2808-05MSD - 1	HR-06-061120MSD	19101	4.3615	-7470A	PPB	15 Jun 2020 15:17:23	U	SMPL
L2997-02 - 1	294	285	0.0247	-7470A	PPB	15 Jun 2020 15:19:29	U	SMPL
CCV20 - 1	CCV20	20887	4.7731	-7470A	PPB	15 Jun 2020 15:21:35	U	SMPL
CCB20 - 1	CCB20	129	-0.0113	-7470A	PPB	15 Jun 2020 15:23:42	U	SMPL
L2999-03 - 1	TRE-1013	261	0.0192	-7470A	PPB	15 Jun 2020 15:25:48	U	SMPL
PB129545TB - 1	PB129542TB	207	0.0067	-7470A	PPB	15 Jun 2020 15:30:18	U	SMPL
L2808-05LX5 - 1		290	0.0258	-7470A	PPB	15 Jun 2020 15:32:23	U	SMPL
L2808-05A - 1		17765	4.0536	-7470A	PPB	15 Jun 2020 15:34:30	U	SMPL
CCV21 - 1	CCV21	20347	4.6487	-7470A	PPB	15 Jun 2020 15:36:36	U	SMPL
CCB21 - 1	CCB21	143	-0.0080	-7470A	PPB	15 Jun 2020 15:38:43	U	SMPL