



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109603

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/15/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	06/15/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	06/15/2020	17:12
4	ICV	1	Water	NA	NA	20.1	6.98	06/15/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.01	06/15/2020	17:16
6	L2808-05	1	Solid	20.03	20	23.8	8.45	06/15/2020	17:19
7	L2815-10	1	Solid	20.04	20	22.5	8.67	06/15/2020	17:22
8	L2997-02	1	Solid	20.02	20	21.8	5.75	06/15/2020	17:25
9	L2998-01	1	Solid	20.03	20	22.5	8.32	06/15/2020	17:26
10	L3013-02	1	Solid	20.02	20	21.5	8.64	06/15/2020	17:30
11	L3013-04	1	Solid	20.03	20	21.4	9.40	06/15/2020	17:32
12	L3013-04DUP	1	Solid	20.02	20	21.5	9.41	06/15/2020	17:33
13	CCV2	1	Water	NA	NA	20.1	12.02	06/15/2020	17:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2997 WorkList ID : 139634 Department : Wet-Chemistry Date : 06-15-2020 13:00:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/11/2020	Solid	L2808-05	Corrosivity	Cool 4 deg C	PSEG05	O61	HR-06-061120	06/11/2020	9045D
06/15/2020	Solid	L2815-10	Corrosivity	Cool 4 deg C	PSEG05		OR-03-061520	06/15/2020	9045D
06/12/2020	Solid	L2997-02	Corrosivity	Cool 4 deg C	PSEG01		294	06/12/2020	9045D
06/12/2020	Solid	L2998-01	Corrosivity	Cool 4 deg C	PSEG01	P22	BS-DRUMS	06/12/2020	9045D
06/15/2020	Solid	L3013-02	Corrosivity	Cool 4 deg C	PSEG01	P23	V-210	06/15/2020	9045D
06/15/2020	Solid	L3013-04	Corrosivity	Cool 4 deg C	PSEG01	P23	V-202	06/15/2020	9045D

Date/Time 06-15-2020 16:00
 Received by: Jp
 Relinquished by: CJP

Date/Time 06-15-2020 17:20
 Received by: CJP
 Relinquished by: Jp