


LB109612

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.4658e-004

C= 2.9573e-002

Rho= 0.9997583

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	-9	0.000	-9				
0.2	0.200	0.210	0.010	730	0.0 %	730				
2.5	2.500	2.550	0.050	10220	0.0 %	10220				
5.0	5.000	4.924	-0.076	19851	0.0 %	19851				
7.5	7.500	7.371	-0.129	29774	0.0 %	29774				
10.0	10.000	10.118	0.118	40913	0.0 %	40913				

LB109612 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Jun 2020 14:02:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-9	-	0.0000	7470A	PPB	16 Jun 2020 14:20:03	S	Std
0.2 - 1	S0.2	730	-	0.2000	7470A	PPB	16 Jun 2020 14:22:10	S	Std
2.5 - 1	S2.5	10220	-	2.5000	7470A	PPB	16 Jun 2020 14:24:17	S	Std
5.0 - 1	S5	19851	-	5.0000	7470A	PPB	16 Jun 2020 14:26:25	S	Std
7.5 - 1	S7.5	29774	-	7.5000	7470A	PPB	16 Jun 2020 14:28:32	S	Std
10.0 - 1	S10	40913	-	10.0000	7470A	PPB	16 Jun 2020 14:30:38	S	Std
ICV86 - 1	ICV86	14947	3.7152	-	7470A	PPB	16 Jun 2020 14:33:16	U	SMPL
ICB86 - 1	ICB86	-93	0.0066	-	7470A	PPB	16 Jun 2020 14:35:21	U	SMPL
CCV25 - 1	CCV25	20215	5.0142	-	7470A	PPB	16 Jun 2020 14:37:27	U	SMPL
CCB25 - 1	CCB25	-124	-0.0010	-	7470A	PPB	16 Jun 2020 14:39:32	U	SMPL
CRA - 1	CRA	716	0.2061	-	7470A	PPB	16 Jun 2020 14:41:38	U	SMPL
HighStd - 1	HighStd	40717	10.0695	-	7470A	PPB	16 Jun 2020 14:43:43	U	SMPL
ChkStd - 1	ChkStd	27872	6.9022	-	7470A	PPB	16 Jun 2020 14:45:50	U	SMPL
PB129577BL - 1	PBW	-169	-0.0121	-	7470A	PPB	16 Jun 2020 14:47:58	U	SMPL
PB129577BS - 1	LCSW	15272	3.7953	-	7470A	PPB	16 Jun 2020 14:50:05	U	SMPL
L2815-10 - 1	OR-03-061520	-96	0.0059	-	7470A	PPB	16 Jun 2020 14:52:12	U	SMPL
L2815-10DUP - 1	OR-03-061520DUP	-37	0.0204	-	7470A	PPB	16 Jun 2020 14:54:19	U	SMPL
L2815-10MS - 1	OR-03-061520MS	15250	3.7899	-	7470A	PPB	16 Jun 2020 14:56:27	U	SMPL
L2815-10MSD - 1	OR-03-061520MSD	14162	3.5216	-	7470A	PPB	16 Jun 2020 14:58:34	U	SMPL
L3013-02 - 1	V-210	-82	0.0094	-	7470A	PPB	16 Jun 2020 15:00:40	U	SMPL
CCV26 - 1	CCV26	20105	4.9870	-	7470A	PPB	16 Jun 2020 15:02:47	U	SMPL
CCB26 - 1	CCB26	-101	0.0047	-	7470A	PPB	16 Jun 2020 15:04:53	U	SMPL
L3013-04 - 1	V-202	-41	0.0195	-	7470A	PPB	16 Jun 2020 15:06:59	U	SMPL
PB129578BL - 1	PBW	-27	0.0229	-	7470A	PPB	16 Jun 2020 15:09:05	U	SMPL
PB129578BS - 1	LCSW	13745	3.4188	-	7470A	PPB	16 Jun 2020 15:11:11	U	SMPL
L2999-01 - 1	31030	35	0.0382	-	7470A	PPB	16 Jun 2020 15:13:18	U	SMPL
L2999-01DUP - 1	31030DUP	57	0.0436	-	7470A	PPB	16 Jun 2020 15:15:25	U	SMPL
L2999-01MS - 1	31030MS	14690	3.6518	-	7470A	PPB	16 Jun 2020 15:20:00	U	SMPL
L2999-01MSD - 1	31030MSD	14989	3.7255	-	7470A	PPB	16 Jun 2020 15:22:07	U	SMPL
PB129579BL - 1	PBW	-87	0.0081	-	7470A	PPB	16 Jun 2020 15:24:14	U	SMPL
PB129579BS - 1	LCSW	14739	3.6639	-	7470A	PPB	16 Jun 2020 15:26:22	U	SMPL
L3014-01 - 1	OILY-WATER	262	0.0942	-	7470A	PPB	16 Jun 2020 15:28:28	U	SMPL
CCV27 - 1	CCV27	20001	4.9614	-	7470A	PPB	16 Jun 2020 15:30:34	U	SMPL
CCB27 - 1	CCB27	-108	0.0029	-	7470A	PPB	16 Jun 2020 15:32:41	U	SMPL
L3014-01DUP - 1	OILY-WATERDUP	278	0.0981	-	7470A	PPB	16 Jun 2020 15:34:47	U	SMPL
L3014-01MS - 1	OILY-WATERMS	11906	2.9653	-	7470A	PPB	16 Jun 2020 15:36:54	U	SMPL
L3014-01MSD - 1	OILY-WATERMSD	11385	2.8369	-	7470A	PPB	16 Jun 2020 15:39:00	U	SMPL
L2815-10LX5 - 1		-123	-0.0008	-	7470A	PPB	16 Jun 2020 15:41:07	U	SMPL
L2815-10A - 1		14711	3.6570	-	7470A	PPB	16 Jun 2020 15:43:15	U	SMPL
L2999-01LX5 - 1		-70	0.0123	-	7470A	PPB	16 Jun 2020 15:45:22	U	SMPL
L2999-01A - 1		14839	3.6886	-	7470A	PPB	16 Jun 2020 15:47:29	U	SMPL
L3014-01LX5 - 1		-53	0.0165	-	7470A	PPB	16 Jun 2020 15:49:37	U	SMPL
L3014-01A - 1		16873	4.1901	-	7470A	PPB	16 Jun 2020 15:51:44	U	SMPL
CCV28 - 1	CCV28	19755	4.9007	-	7470A	PPB	16 Jun 2020 15:53:51	U	SMPL
CCB28 - 1	CCB28	-118	0.0005	-	7470A	PPB	16 Jun 2020 15:55:57	U	SMPL
PB129577TB - 1	PB129549TB	-43	0.0190	-	7470A	PPB	16 Jun 2020 15:58:04	U	SMPL
PB129578TB - 1	PB129550TB	68	0.0463	-	7470A	PPB	16 Jun 2020 16:00:10	U	SMPL
CCV29 - 1	CCV29	20143	4.9964	-	7470A	PPB	16 Jun 2020 16:02:17	U	SMPL
CCB29 - 1	CCB29	-112	0.0020	-	7470A	PPB	16 Jun 2020 16:04:23	U	SMPL