

LB109629


7470A

INSTRUMENT 19: CY1

Linear

 μ Abs.:

36603



A= 0.0000e+000

B= 2.7058e-004

C= -2.0282e-003

Rho= 0.9998456

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-18	0.000	-18				
0.2	0.200	0.184	-0.016	688	0.0 %	688				
2.5	2.500	2.460	-0.040	9100	0.0 %	9100				
5.0	5.000	5.059	0.059	18703	0.0 %	18703				
7.5	7.500	7.602	0.102	28101	0.0 %	28101				
10.0	10.000	9.902	-0.098	36603	0.0 %	36603				

LB109629 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 17 Jun 2020 14:04:19				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-18	-	0.00007470A	PPB	17 Jun 2020 14:07:17	S	Std
0.2 - 1	50.2	688	-	0.20007470A	PPB	17 Jun 2020 14:09:24	S	Std
2.5 - 1	52.5	9100	-	2.50007470A	PPB	17 Jun 2020 14:11:32	S	Std
5.0 - 1	55	18703	-	5.00007470A	PPB	17 Jun 2020 14:13:39	S	Std
7.5 - 1	57.5	28101	-	7.50007470A	PPB	17 Jun 2020 14:15:46	S	Std
10.0 - 1	510	36603	-	10.00007470A	PPB	17 Jun 2020 14:20:21	S	Std
ICV88 - 1	ICV88	15797	4.2724	-7470A	PPB	17 Jun 2020 14:22:53	U	SMPL
ICB88 - 1	ICB88	-122	-0.0350	-7470A	PPB	17 Jun 2020 14:24:58	U	SMPL
CCV33 - 1	CCV33	18367	4.9678	-7470A	PPB	17 Jun 2020 14:27:04	U	SMPL
CCB33 - 1	CCB33	-114	-0.0329	-7470A	PPB	17 Jun 2020 14:29:10	U	SMPL
CRA - 1	CRA	645	0.1725	-7470A	PPB	17 Jun 2020 14:31:15	U	SMPL
HighStd - 1	HighStd	36581	9.8962	-7470A	PPB	17 Jun 2020 14:33:21	U	SMPL
ChkStd - 1	ChkStd	26327	7.1216	-7470A	PPB	17 Jun 2020 14:35:27	U	SMPL
PB129594BL - 1	PBW	-160	-0.0453	-7470A	PPB	17 Jun 2020 14:37:35	U	SMPL
PB129594BS - 1	LCSW	14598	3.9480	-7470A	PPB	17 Jun 2020 14:39:42	U	SMPL
L2890-04 - 1	72-11979-72-11969	-106	-0.0307	-7470A	PPB	17 Jun 2020 14:41:49	U	SMPL
L2890-04DUP - 1	72-11979-72-11969DUP	-44	-0.0139	-7470A	PPB	17 Jun 2020 14:43:56	U	SMPL
L2890-04MS - 1	72-11979-72-11969MS	14053	3.8005	-7470A	PPB	17 Jun 2020 14:46:04	U	SMPL
L2890-04MSD - 1	72-11979-72-11969MSD	15700	4.2461	-7470A	PPB	17 Jun 2020 14:48:12	U	SMPL
L2890-06 - 1	CHRT-24245	317	0.0837	-7470A	PPB	17 Jun 2020 14:50:18	U	SMPL
CCV34 - 1	CCV34	18548	5.0168	-7470A	PPB	17 Jun 2020 14:52:23	U	SMPL
CCB34 - 1	CCB34	-105	-0.0304	-7470A	PPB	17 Jun 2020 14:54:29	U	SMPL
L2890-08 - 1	26572-25062-251686	-7	-0.0039	-7470A	PPB	17 Jun 2020 14:56:36	U	SMPL
L2890-04LX5 - 1		-29	-0.0099	-7470A	PPB	17 Jun 2020 14:58:43	U	SMPL
L2890-04A - 1		15570	4.2110	-7470A	PPB	17 Jun 2020 15:00:49	U	SMPL
CCV35 - 1	CCV35	18754	5.0725	-7470A	PPB	17 Jun 2020 15:05:02	U	SMPL
CCB35 - 1	CCB35	-103	-0.0299	-7470A	PPB	17 Jun 2020 15:07:09	U	SMPL