



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109634

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/17/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.01	06/17/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	06/17/2020	17:32
4	ICV	1	Water	NA	NA	20.1	7.00	06/17/2020	17:33
5	CCV1	1	Water	NA	NA	20.2	2.01	06/17/2020	17:35
6	L3048-02	1	Water	NA	NA	20.1	7.98	06/17/2020	18:20
7	L3048-02DUP	1	Water	NA	NA	20.2	7.99	06/17/2020	18:21
8	CCV2	1	Water	NA	NA	20.2	12.02	06/17/2020	18:22

HP 109634

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3248 W WorkList ID : 139699 Department : Wet-Chemistry Date : 06-17-2020 15:41:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2020	Water	L3048-02-F	Corrosivity	Cool 4 deg C	SEAC01	P42	GROUNDWATER	06/17/2020	9040C

Date/Time 06-17-2020 15:50
Received by: JS
Relinquished by: KS

Date/Time 06-17-2020 17:20
Received by: KS
Relinquished by: JS