



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109638

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/15/2020	17:20
2	CAL2	1	Water	NA	NA	20.1	7.01	06/15/2020	17:21
3	CAL3	1	Water	NA	NA	20.1	10.03	06/15/2020	17:22
4	ICV	1	Water	NA	NA	20.1	7.00	06/15/2020	17:25
5	CCV1	1	Water	NA	NA	20.2	2.01	06/15/2020	17:26
6	L3014-01	1	Water	NA	NA	16.4	7.64	06/15/2020	17:29
7	L3014-01DUP	1	Water	NA	NA	16.5	7.65	06/15/2020	17:30
8	CCV2	1	Water	NA	NA	20.1	12.02	06/15/2020	17:32

WORKLIST(Hardcopy Internal Chain)

YB 109638

WorkList Name : PH L3014

WorkList ID : 139726

Department : Wet-Chemistry

Date : 06-14-2020 17:10:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2020	Water	L-3014-01	pH	Cool 4 deg C	PSEG01	P23	OILY-WATER	06/15/2020	9040C

Date/Time 06-14-2020 17:15
 Received by: JS
 Relinquished by: JS

Date/Time 06-14-2020 18:25
 Received by: JS
 Relinquished by: JS