



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109639

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/18/2020	11:30
2	CAL2	1	Water	NA	NA	20.2	7.01	06/18/2020	11:31
3	CAL3	1	Water	NA	NA	20.2	10.03	06/18/2020	11:32
4	ICV	1	Water	NA	NA	20.2	7.00	06/18/2020	11:35
5	CCV1	1	Water	NA	NA	20.2	2.01	06/18/2020	11:36
6	L3035-01	1	Solid	20.03	20	20.7	7.10	06/18/2020	11:39
7	L3035-01DUP	1	Solid	20.04	20	20.7	7.11	06/18/2020	11:40
8	L3035-05	1	Solid	20.02	20	20.8	8.03	06/18/2020	11:41
9	L3035-09	1	Solid	20.05	20	21.1	8.78	06/18/2020	11:42
10	L3035-13	1	Solid	20.03	20	21.1	9.00	06/18/2020	11:44
11	CCV2	1	Water	NA	NA	20.1	12.02	06/18/2020	11:45

WORKLIST(Hardcopy Internal Chain)

12109639

WorkList Name : PH L3035

WorkList ID : 139725

Department : Wet-Chemistry

Date : 06-18-2020 11:21:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2020	Solid	L3035-01	pH	Cool 4 deg C	PSEG01	P43	1-JUNIVERSITY-SW-5	06/17/2020	9045D
06/17/2020	Solid	L3035-05	pH	Cool 4 deg C	PSEG01	P43	2-JUNIVERSITY-SW-3	06/17/2020	9045D
06/17/2020	Solid	L3035-09	pH	Cool 4 deg C	PSEG01	P43	1-SP-3-5	06/17/2020	9045D
06/17/2020	Solid	L3035-13	pH	Cool 4 deg C	PSEG01	P43	2-SP-3-5	06/17/2020	9045D

Date/Time 06-18-2020 11:25
 Received by: JA KS
 Relinquished by:

Date/Time 06-18-2020 12:40
 Received by: KS
 Relinquished by: Jn