



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109640

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/18/2020	11:30
2	CAL2	1	Water	NA	NA	20.2	7.01	06/18/2020	11:31
3	CAL3	1	Water	NA	NA	20.2	10.03	06/18/2020	11:32
4	ICV	1	Water	NA	NA	20.2	7.00	06/18/2020	11:35
5	CCV1	1	Water	NA	NA	20.2	2.01	06/18/2020	11:36
6	L2960-01	1	Water	NA	NA	21.5	7.48	06/18/2020	11:50
7	L2960-01DUP	1	Water	NA	NA	21.6	7.49	06/18/2020	11:51
8	L3057-01	1	Water	NA	NA	21.4	7.24	06/18/2020	12:00
9	L3057-02	1	Water	NA	NA	21.5	7.02	06/18/2020	12:02
10	L3057-03	1	Water	NA	NA	21.0	6.84	06/18/2020	12:05
11	CCV2	1	Water	NA	NA	20.1	12.02	06/18/2020	12:10

LB 109640

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3057 WorkList ID : 139728 Department : Wet-Chemistry Date : 06-18-2020 11:26:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2020	Water	L2960-01 - A	pH	Cool 4 deg C	STER17	O11	JS9600	06/04/2020	SM 4500-H B
06/17/2020	Water	L3057-01 - A	pH	Cool 4 deg C	STER07	P61	CELLS-A	06/17/2020	SM 4500-H B
06/17/2020	Water	L3057-02 - A	pH	Cool 4 deg C	STER07	P61	CELLS-B/C	06/17/2020	SM 4500-H B
06/17/2020	Water	L3057-03 - A	pH	Cool 4 deg C	STER07	P61	CELLS-D	06/17/2020	SM 4500-H B

Date/Time 06-18-2020 11:30
 Received by: J
 Relinquished by: AM

Date/Time 06-18-2020 11:30
 Received by: AM
 Relinquished by: J