

Analysis Method: SM2510 BANALYST: AlexanderParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB109642Meter ID: WC pH Meter-1Thermometer ID: pH MeterTempProbe

Reagent/Standard	Lot/Log #
conductivity standard	WP81395

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB109642BL		1	20.6	0.72	0.79	06/18/2020	12:48
2	LB109642BS	1413	1	20.0	1276.00	1410.72	06/18/2020	12:52
3	L2960-01		1	20.2	1.52	1.67	06/18/2020	12:55
4	L3020-15		1	20.1	1002.00	1105.46	06/18/2020	12:58
5	L3020-20		1	20.1	1008.00	1112.08	06/18/2020	13:02
6	L3020-23		1	19.9	790.00	875.26	06/18/2020	13:05
7	L3020-23DUP		1	19.9	793.00	878.58	06/18/2020	13:08

Note : Conductivity result is corrected to 25°C

Result = <u> </u> (Reading) 1 + 0.0191 (Temperature - 25)

LB109642
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Conductance-61820 WorkList ID : 139721 Department : Wet-Chemistry Date : 06-18-2020 09:48:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2020	Water	L2960-01 C	Conductance	Cool 4 deg C	STER17	O11	JS9600	06/04/2020	SM2510 B
06/29/2020	Water	L3020-15 C	Conductance	Cool 4 deg C	CHAZ02	P43	MW-26	06/17/2020	SM2510 B
06/30/2020	Water	L3020-20 C	Conductance	Cool 4 deg C	CHAZ02	P43	MW-23	06/17/2020	SM2510 B
06/30/2020	Water	L3020-23 A	Conductance	Cool 4 deg C	CHAZ02	P43	MW-27	06/17/2020	SM2510 B

Date/Time 6/18/2020 12:15
 Received by: AM
 Relinquished by: JA

Date/Time 6/18/2020 15:10
 Received by: JD
 Relinquished by: AM