



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB109643

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP84901
Chlorine Calibration std, 0.1ppm	WP84896
Chlorine Calibration std, 0.2ppm	WP84897
Chlorine Calibration std, 0.8ppm	WP84899
Chlorine Calibration std, 0.0ppm	WP84895
Chlorine Calibration std, 1.6ppm	WP84900
Residual Chlorine Calibration and CCV std, 0	WP84898

Intercept: 0.0046

Slope: 1.0589

Regression: 0.999981

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.010	0.010	06/18/2020	13:35
2	CAL2	0.1	1	0.000	0.110	0.110	06/18/2020	13:38
3	CAL3	0.2	1	0.000	0.210	0.210	06/18/2020	13:40
4	CAL4	0.4	1	0.000	0.430	0.430	06/18/2020	13:42
5	CAL5	0.8	1	0.000	0.850	0.850	06/18/2020	13:45
6	CAL6	1.6	1	0.000	1.700	1.700	06/18/2020	13:47



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB109643

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4500	0.4500	0.4210	06/18/2020	13:50
2	ICB				1	0.0000	0.0000	0.0000	-0.0040	06/18/2020	13:53
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.3920	06/18/2020	13:55
4	CCB1				1	0.0000	0.0000	0.0000	-0.0040	06/18/2020	13:58
5	LB109643BL	50	50		1	0.0000	0.0000	0.0000	-0.0040	06/18/2020	14:01
6	LB109643BS	50	50	0.4	1	0.0000	0.4500	0.4500	0.4210	06/18/2020	14:03
7	L2960-01	50	50		1	0.0000	0.0100	0.0100	0.0050	06/18/2020	14:05
8	L2960-01DUP	50	50		1	0.0000	0.0100	0.0100	0.0050	06/18/2020	14:07
9	CCV2			0.4	1	0.0000	0.4400	0.4400	0.4110	06/18/2020	14:10
10	CCB2				1	0.0000	0.0000	0.0000	-0.0040	06/18/2020	14:12

LB109643 WORKLIST(Hardcopy Internal Chain)

WorkList Name : ResidualChlorine-61820 WorkList ID : 139724 Department : Wet-Chemistry Date : 06-18-2020 11:09:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2020	Water	L2960-01 A	Residual Chlorine	Cool 4 deg C	STER17	O11	JS9600	06/04/2020	SM4500 Cl G

Date/Time 6/18/2020 12:15
 Received by: Am
 Relinquished by: Am

Date/Time 6/18/2020 15:10
 Received by: JN
 Relinquished by: Am