



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109646

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/18/2020	12:35
2	CAL2	1	Water	NA	NA	20.1	7.00	06/18/2020	12:36
3	CAL3	1	Water	NA	NA	20.1	10.03	06/18/2020	12:37
4	ICV	1	Water	NA	NA	20.2	7.01	06/18/2020	12:39
5	CCV1	1	Water	NA	NA	20.1	2.02	06/18/2020	12:40
6	L3056-01	1	Solid	20.03	20	20.4	8.03	06/18/2020	12:41
7	L3056-01DUP	1	Solid	20.04	20	20.5	8.04	06/18/2020	12:42
8	L3056-02	1	Solid	20.03	20	20.6	8.51	06/18/2020	12:44
9	CCV2	1	Water	NA	NA	20.2	12.02	06/18/2020	12:45

HB109646

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3056 WorkList ID : 139733 Department : Wet-Chemistry Date : 06-18-2020 13:02:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2020	Solid	L3056-01- <u>CS</u>	pH	Cool 4 deg C	JPCL01	P61	SAMPLE-A-LOT-11	06/17/2020	9045D
06/17/2020	Solid	L3056-02- <u>CS</u>	pH	Cool 4 deg C	JPCL01	P61	SAMPLE-B-LOT-11	06/17/2020	9045D

Date/Time 06-18-2020 12:30
Received by: JP
Relinquished by: JP

Date/Time 06-18-2020 14:00
Received by: JP
Relinquished by: JP