



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109651

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/18/2020	14:15
2	CAL2	1	Water	NA	NA	20.1	7.00	06/18/2020	14:16
3	CAL3	1	Water	NA	NA	20.1	10.03	06/18/2020	14:19
4	ICV	1	Water	NA	NA	20.1	6.98	06/18/2020	14:20
5	CCV1	1	Water	NA	NA	20.1	2.01	06/18/2020	14:21
6	L3054-01	1	Water	NA	NA	17.8	5.54	06/18/2020	14:25
7	L3054-01DUP	1	Water	NA	NA	17.9	5.55	06/18/2020	14:26
8	CCV2	1	Water	NA	NA	20.1	12.02	06/18/2020	14:29

VB 109651

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3054 WorkList ID : 139739 Department : Wet-Chemistry Date : 06-18-2020 14:01:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2020	Water	L3054-01- <u>E</u>	pH	Cool 4 deg C	PSEG01	P52	218	06/18/2020	9040C

Date/Time 06-18-2020 14:10
 Received by: Jp
 Relinquished by: Jp

Date/Time 06-18-2020 15:15
 Received by: Jp
 Relinquished by: Jp