

L3109652


7470A

INSTRUMENT ID: CY4

Linear

μ Abs.:

38689



A= 0.0000e+000

B= 2.5989e-004

C= 2.3140e-003

Rho= 0.9994862

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	-16	0.000	-16				
0.2	0.200	0.195	-0.005	740	0.0 %	740				
2.5	2.500	2.459	-0.041	9452	0.0 %	9452				
5.0	5.000	5.194	0.194	19978	0.0 %	19978				
7.5	7.500	7.297	-0.203	28069	0.0 %	28069				
10.0	10.000	10.057	0.057	38689	0.0 %	38689				

LB109652 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 18 Jun 2020 14:52:14							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-16	-	0.00007470A	PPB		18 Jun 2020 14:57:25	S	Std
0.2 - 1	50.2	740	-	0.20007470A	PPB		18 Jun 2020 14:59:33	S	Std
2.5 - 1	52.5	9452	-	2.50007470A	PPB		18 Jun 2020 15:01:41	S	Std
5.0 - 1	55	19978	-	5.00007470A	PPB		18 Jun 2020 15:03:49	S	Std
7.5 - 1	57.5	28069	-	7.50007470A	PPB		18 Jun 2020 15:05:56	S	Std
10.0 - 1	59	38689	-	10.00007470A	PPB		18 Jun 2020 15:10:35	S	Std
ICV90 - 1	ICV90	14481	3.7657	-7470A	PPB		18 Jun 2020 15:13:06	U	SMPL
ICB90 - 1	ICB90	-112	-0.0268	-7470A	PPB		18 Jun 2020 15:15:11	U	SMPL
CCV42 - 1	CCV42	20015	5.2039	-7470A	PPB		18 Jun 2020 15:17:17	U	SMPL
CCB42 - 1	CCB42	-129	-0.0312	-7470A	PPB		18 Jun 2020 15:19:23	U	SMPL
CRA - 1	CRA	717	0.1887	-7470A	PPB		18 Jun 2020 15:21:28	U	SMPL
HighStd - 1	HighStd	38282	9.9513	-7470A	PPB		18 Jun 2020 15:23:34	U	SMPL
ChkStd - 1	ChkStd	27399	7.1229	-7470A	PPB		18 Jun 2020 15:25:40	U	SMPL
PB129623BL - 1	PBW	-143	-0.0348	-7470A	PPB		18 Jun 2020 15:27:48	U	SMPL
PB129623BS - 1	LCSW	14651	3.8099	-7470A	PPB		18 Jun 2020 15:29:55	U	SMPL
L2809-04 - 1	HD-01-061720	-91	-0.0213	-7470A	PPB		18 Jun 2020 15:32:02	U	SMPL
L2809-04DUP - 1	HD-01-061720DUP	-41	-0.0083	-7470A	PPB		18 Jun 2020 15:34:09	U	SMPL
L2809-04MS - 1	HD-01-061720MS	14217	3.6971	-7470A	PPB		18 Jun 2020 15:36:17	U	SMPL
L2809-04MSD - 1	HD-01-061720MSD	15059	3.9159	-7470A	PPB		18 Jun 2020 15:38:24	U	SMPL
L2811-04 - 1	NB-08-061620	-112	-0.0268	-7470A	PPB		18 Jun 2020 15:40:30	U	SMPL
CCV43 - 1	CCV43	19896	5.1730	-7470A	PPB		18 Jun 2020 15:42:36	U	SMPL
CCB43 - 1	CCB43	-114	-0.0273	-7470A	PPB		18 Jun 2020 15:44:42	U	SMPL
L2812-10 - 1	OK-03-061720	-45	-0.0094	-7470A	PPB		18 Jun 2020 15:46:48	U	SMPL
L2813-12 - 1	EO-01-061720	-65	-0.0146	-7470A	PPB		18 Jun 2020 15:48:55	U	SMPL
L2817-10 - 1	SU-04-061720	-70	-0.0159	-7470A	PPB		18 Jun 2020 15:51:01	U	SMPL
L2818-04 - 1	TR-05-061720	-30	-0.0055	-7470A	PPB		18 Jun 2020 15:53:08	U	SMPL
L3024-02 - 1	RO-21	-29	-0.0052	-7470A	PPB		18 Jun 2020 15:55:14	U	SMPL
L3024-05 - 1	RO-25	-18	-0.0024	-7470A	PPB		18 Jun 2020 15:57:21	U	SMPL
L3024-08 - 1	RO-16	-23	-0.0037	-7470A	PPB		18 Jun 2020 15:59:29	U	SMPL
L3024-11 - 1	RO-7	8	0.0044	-7470A	PPB		18 Jun 2020 16:01:36	U	SMPL
L3024-14 - 1	RO-1	-15	-0.0016	-7470A	PPB		18 Jun 2020 16:03:44	U	SMPL
L3024-17 - 1	RO-6	-34	-0.0065	-7470A	PPB		18 Jun 2020 16:05:50	U	SMPL
CCV44 - 1	CCV44	19582	5.0914	-7470A	PPB		18 Jun 2020 16:07:57	U	SMPL
CCB44 - 1	CCB44	-111	-0.0265	-7470A	PPB		18 Jun 2020 16:10:03	U	SMPL
L3024-20 - 1	RO-17	-19	-0.0026	-7470A	PPB		18 Jun 2020 16:12:10	U	SMPL
L3024-23 - 1	RO-11	-28	-0.0050	-7470A	PPB		18 Jun 2020 16:14:16	U	SMPL
L3046-13 - 1	ISS-A-C	-26	-0.0044	-7470A	PPB		18 Jun 2020 16:16:23	U	SMPL
L3046-14 - 1	ISS-B-C	-33	-0.0063	-7470A	PPB		18 Jun 2020 16:18:29	U	SMPL
L3046-15 - 1	ISS-C-C	-21	-0.0031	-7470A	PPB		18 Jun 2020 16:20:36	U	SMPL
L3048-01 - 1	SOIL-CUTTINGS	-16	-0.0018	-7470A	PPB		18 Jun 2020 16:22:43	U	SMPL
PB129624BL - 1	PBW	-43	-0.0089	-7470A	PPB		18 Jun 2020 16:24:50	U	SMPL
PB129624BS - 1	LCSW	14556	3.7852	-7470A	PPB		18 Jun 2020 16:26:58	U	SMPL
L3048-02 - 1	GROUNDWATER	17862	4.6444	-7470A	PPB		18 Jun 2020 16:29:05	U	SMPL
L3048-02DUP - 1	GROUNDWATERDUP	17711	4.6052	-7470A	PPB		18 Jun 2020 16:31:12	U	SMPL
CCV45 - 1	CCV45	19543	5.0813	-7470A	PPB		18 Jun 2020 16:33:19	U	SMPL
CCB45 - 1	CCB45	-100	-0.0237	-7470A	PPB		18 Jun 2020 16:35:27	U	SMPL
L3048-02MS - 1	GROUNDWATERMS	36223	9.4162	-7470A	PPB		18 Jun 2020 16:37:33	U	SMPL
L3048-02MSD - 1	GROUNDWATERMSD	33274	8.6498	-7470A	PPB		18 Jun 2020 16:39:39	U	SMPL
PB129623TB - 1	PB129584TB	-170	-0.0419	-7470A	PPB		18 Jun 2020 16:41:48	U	SMPL
L2809-04LX5 - 1		-45	-0.0094	-7470A	PPB		18 Jun 2020 16:43:55	U	SMPL
L2809-04A - 1		14425	3.7512	-7470A	PPB		18 Jun 2020 16:46:01	U	SMPL
L3048-02LX5 - 1		2797	0.7292	-7470A	PPB		18 Jun 2020 16:48:08	U	SMPL
L3048-02A - 1		31054	8.0728	-7470A	PPB		18 Jun 2020 16:50:16	U	SMPL
CCV46 - 1	CCV46	19743	5.1333	-7470A	PPB		18 Jun 2020 16:52:23	U	SMPL
CCB46 - 1	CCB46	-114	-0.0273	-7470A	PPB		18 Jun 2020 16:54:32	U	SMPL