


LB109657

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	11	0.000	11				
0.2	0.200	0.214	0.014	1215	0.0 %	1215				
2.5	2.500	2.464	-0.036	14672	0.0 %	14672				
5.0	5.000	5.010	0.010	29900	0.0 %	29900				
7.5	7.500	7.480	-0.020	44676	0.0 %	44676				
10.0	10.000	10.018	0.018	59858	0.0 %	59858				

LB109657 INSTRUMENT ID : CV1

Method: 7471B	Operator: Admin	Date of Analysis: 19 Jun 2020 08:41:28							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.0000	7471B	PPB	19 Jun 2020 08:44:57	S	Std
0.2 - 1	50.2	1215	-	0.2000	7471B	PPB	19 Jun 2020 08:47:14	S	Std
2.5 - 1	52.5	14672	-	2.5000	7471B	PPB	19 Jun 2020 08:49:31	S	Std
5.0 - 1	55	29900	-	5.0000	7471B	PPB	19 Jun 2020 08:51:48	S	Std
7.5 - 1	57.5	44676	-	7.5000	7471B	PPB	19 Jun 2020 08:54:05	S	Std
10.0 - 1	58	59858	-	10.0000	7471B	PPB	19 Jun 2020 08:59:02	S	Std
ICV91 - 1	ICV91	23906	4.0079	-	7471B	PPB	19 Jun 2020 09:01:44	U	SMPL
ICB91 - 1	ICB91	-104	-0.0061	-	7471B	PPB	19 Jun 2020 09:03:59	U	SMPL
CCV47 - 1	CCV47	29897	5.0094	-	7471B	PPB	19 Jun 2020 09:06:15	U	SMPL
CCB47 - 1	CCB47	-98	-0.0051	-	7471B	PPB	19 Jun 2020 09:08:30	U	SMPL
CRA - 1	CRA	1208	0.2132	-	7471B	PPB	19 Jun 2020 09:10:46	U	SMPL
HighStd - 1	HighStd	59654	9.9842	-	7471B	PPB	19 Jun 2020 09:13:02	U	SMPL
ChkStd - 1	ChkStd	40910	6.8506	-	7471B	PPB	19 Jun 2020 09:15:17	U	SMPL
PB129634BL - 1	PBS	-192	-0.0209	-	7471B	PPB	19 Jun 2020 09:17:34	U	SMPL
PB129634BS - 1	LCSS	20859	3.4985	-	7471B	PPB	19 Jun 2020 09:19:52	U	SMPL
L3035-01 - 1	1-UNIVERSITY-SW-5	2056	0.3550	-	7471B	PPB	19 Jun 2020 09:22:09	U	SMPL
L3035-05 - 1	2-UNIVERISTY-SW-3	2417	0.4153	-	7471B	PPB	19 Jun 2020 09:24:25	U	SMPL
L3035-09 - 1	1-SP-3-5	2283	0.3929	-	7471B	PPB	19 Jun 2020 09:26:42	U	SMPL
L3035-13 - 1	2-SP-3-5	2854	0.4884	-	7471B	PPB	19 Jun 2020 09:28:59	U	SMPL
L3052-02 - 1	IWS	1953	0.3377	-	7471B	PPB	19 Jun 2020 09:31:13	U	SMPL
CCV48 - 1	CCV48	29986	5.0243	-	7471B	PPB	19 Jun 2020 09:33:29	U	SMPL
CCB48 - 1	CCB48	-107	-0.0067	-	7471B	PPB	19 Jun 2020 09:35:44	U	SMPL
L3052-02DUP - 1	IWSDUP	1868	0.3235	-	7471B	PPB	19 Jun 2020 09:38:00	U	SMPL
L3052-02MS - 1	IWSMS	22214	3.7250	-	7471B	PPB	19 Jun 2020 09:40:15	U	SMPL
L3052-02MSD - 1	IWSMSD	22867	3.8342	-	7471B	PPB	19 Jun 2020 09:42:31	U	SMPL
L3055-01 - 1	AMITY-ST-ENTRANCE	354	0.0704	-	7471B	PPB	19 Jun 2020 09:44:47	U	SMPL
L3055-02 - 1	CENTER	7499	1.2649	-	7471B	PPB	19 Jun 2020 09:47:03	U	SMPL
L3055-03 - 1	FAIRMOUNT-REAR	11659	1.9604	-	7471B	PPB	19 Jun 2020 09:49:19	U	SMPL
L3056-01 - 1	SAMPLE-A-LOT-11	47824	8.0065	-	7471B	PPB	19 Jun 2020 09:51:35	U	SMPL
L3056-02 - 1	SAMPLE-B-LOT-11	13462	2.2618	-	7471B	PPB	19 Jun 2020 09:53:52	U	SMPL
L3052-02LX5 - 1		314	0.0637	-	7471B	PPB	19 Jun 2020 09:56:11	U	SMPL
L3052-02A - 1		24100	4.0403	-	7471B	PPB	19 Jun 2020 09:58:27	U	SMPL
CCV49 - 1	CCV49	30065	5.0375	-	7471B	PPB	19 Jun 2020 10:00:42	U	SMPL
CCB49 - 1	CCB49	-109	-0.0070	-	7471B	PPB	19 Jun 2020 10:02:58	U	SMPL