



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109658

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/19/2020	10:05
2	L3054-01		1	8	100.00	>212.0	>212.0	06/19/2020	10:35
3	L3054-01DUP		1	8	100.00	>212.0	>212.0	06/19/2020	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp61920 WorkList ID : 139761 Department : Wet-Chemistry Date : 06-19-2020 09:35:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/25/2020	Water	L3054-01	Flash Point	Cool 4 deg C	PSEG01	P52	218	06/18/2020	1010A

Date/Time 06/19/2020 09:40
 Received by: KS
 Relinquished by: JS

Date/Time 06/19/2020 11:15
 Received by: JP
 Relinquished by: KS