


LB109665

ISM02.4

INSTRUMENT ID: CV4



Linear

A= 0.0000e+000

B= 2.3192e-004

C= 1.1727e-002

Rho= 0.9997591

Accept=Accepted

Slope
 $\gamma = 1.9715 \text{ except}$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	-76	0.000	-76					=
0.05	0.050					223					4
0.20	0.200	0.207	0.007	844	0.0 %	844					4
2.50	2.500	2.612	0.112	11211	0.0 %	11211					4
5.00	5.000	4.856	-0.144	20888	0.0 %	20888					-3
7.50	7.500	7.462	-0.038	32125	0.0 %	32125					-1
10.0	10.000	10.068	0.068	43361	0.0 %	43361					1

LB109665
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-76	-		0 ISM02.4	PPB	6/19/2020 13:32	Std
	0.2 S0.2	844	-		0.2 ISM02.4	PPB	6/19/2020 13:34	Std
	2.5 S2.5	11211	-		2.5 ISM02.4	PPB	6/19/2020 13:37	Std
	5 S5.0	20888	-		5 ISM02.4	PPB	6/19/2020 13:39	Std
	7.5 S7.5	32125	-		7.5 ISM02.4	PPB	6/19/2020 13:46	Std
	10 S10.0	43361	-		10 ISM02.4	PPB	6/19/2020 13:51	Std
ICV	ICV	17905	4.1643 -		ISM02.4	PPB	6/19/2020 13:54	SMPL
ICB	ICB	-172	-0.0282 -		ISM02.4	PPB	6/19/2020 13:56	SMPL
CCV050	CCV050	20639	4.7984 -		ISM02.4	PPB	6/19/2020 13:58	SMPL
CCB050	CCB050	-187	-0.0316 -		ISM02.4	PPB	6/19/2020 14:01	SMPL
PB129650BL	PBW650	-98	-0.011 -		ISM02.4	PPB	6/19/2020 14:03	SMPL
L2981-01	MBFSY4	-85	-0.008 -		ISM02.4	PPB	6/19/2020 14:05	SMPL
L2981-02	MBFSY5	-92	-0.0096 -		ISM02.4	PPB	6/19/2020 14:07	SMPL
L2981-03	MBFSY6	-89	-0.0089 -		ISM02.4	PPB	6/19/2020 14:10	SMPL
L2981-04	MBFT08	-103	-0.0122 -		ISM02.4	PPB	6/19/2020 14:12	SMPL
L2981-05	MBFSY6D	-98	-0.011 -		ISM02.4	PPB	6/19/2020 14:14	SMPL
L2981-06	MBFSY6S	4624	1.0841 -		ISM02.4	PPB	6/19/2020 14:17	SMPL
CCV051	CCV051	20710	4.8149 -		ISM02.4	PPB	6/19/2020 14:19	SMPL
CCB051	CCB051	-159	-0.0251 -		ISM02.4	PPB	6/19/2020 14:21	SMPL