

LB109669


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40840



A= 0.0000e+000

B= 2.4365e-004

C= 3.3666e-002

Rho= 0.9996818

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	-1	0.000	-1				
0.2	0.200	0.220	0.020	764	0.0 %	764				
2.5	2.500	2.515	0.015	10182	0.0 %	10182				
5.0	5.000	4.818	-0.182	19636	0.0 %	19636				
7.5	7.500	7.630	0.130	31177	0.0 %	31177				
10.0	10.000	9.984	-0.016	40840	0.0 %	40840				

LB109669 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 19 Jun 2020 14:44:13						
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-1	-	0.00007470A	PPB	19 Jun 2020 14:46:00	S	Std
0.2 - 1	50.2	764	-	0.20007470A	PPB	19 Jun 2020 14:48:08	S	Std
2.5 - 1	52.5	10182	-	2.50007470A	PPB	19 Jun 2020 14:50:16	S	Std
5.0 - 1	55	19636	-	5.00007470A	PPB	19 Jun 2020 14:52:23	S	Std
7.5 - 1	57.5	31177	-	7.50007470A	PPB	19 Jun 2020 14:54:30	S	Std
10.0 - 1	510	40840	-	10.00007470A	PPB	19 Jun 2020 14:56:36	S	Std
ICV92 - 1	ICV92	15219	3.7418	-7470A	PPB	19 Jun 2020 14:59:10	U	SMPL
ICB92 - 1	ICB92	-84	0.0132	-7470A	PPB	19 Jun 2020 15:01:15	U	SMPL
CCV52 - 1	CCV52	19272	4.7293	-7470A	PPB	19 Jun 2020 15:03:20	U	SMPL
CCB52 - 1	CCB52	-70	0.0166	-7470A	PPB	19 Jun 2020 15:05:26	U	SMPL
CRA - 1	CRA	741	0.2142	-7470A	PPB	19 Jun 2020 15:07:32	U	SMPL
HighStd - 1	HighStd	39348	9.6208	-7470A	PPB	19 Jun 2020 15:09:38	U	SMPL
ChkStd - 1	ChkStd	26687	6.5360	-7470A	PPB	19 Jun 2020 15:11:43	U	SMPL
PB129651BL - 1	PBW	-30	0.0264	-7470A	PPB	19 Jun 2020 15:18:34	U	SMPL
PB129651BS - 1	LCSW	14768	3.6319	-7470A	PPB	19 Jun 2020 15:20:40	U	SMPL
L3034-03 - 1	TP-1	-79	0.0144	-7470A	PPB	19 Jun 2020 15:22:47	U	SMPL
L3034-03DUP - 1	TP-1DUP	-45	0.0227	-7470A	PPB	19 Jun 2020 15:24:54	U	SMPL
L3034-03MS - 1	TP-1MS	15661	3.8495	-7470A	PPB	19 Jun 2020 15:27:01	U	SMPL
L3034-03MSD - 1	TP-1MSD	15153	3.7257	-7470A	PPB	19 Jun 2020 15:29:09	U	SMPL
L3034-06 - 1	TP-2	-87	0.0125	-7470A	PPB	19 Jun 2020 15:31:15	U	SMPL
CCV53 - 1	CCV53	19198	4.7113	-7470A	PPB	19 Jun 2020 15:33:21	U	SMPL
CCB53 - 1	CCB53	-82	0.0137	-7470A	PPB	19 Jun 2020 15:35:26	U	SMPL
L3034-09 - 1	TP-3	-11	0.0310	-7470A	PPB	19 Jun 2020 15:37:32	U	SMPL
L3034-12 - 1	TP-4	-15	0.0300	-7470A	PPB	19 Jun 2020 15:39:38	U	SMPL
L3034-15 - 1	TP-5	-32	0.0259	-7470A	PPB	19 Jun 2020 15:41:45	U	SMPL
L3035-02 - 1	1-UNIVERSITY-SW-5	-22	0.0283	-7470A	PPB	19 Jun 2020 15:43:52	U	SMPL
L3035-06 - 1	2-UNIVERSITY-SW-3	-33	0.0256	-7470A	PPB	19 Jun 2020 15:45:58	U	SMPL
L3035-10 - 1	1-SP-3-5	-21	0.0285	-7470A	PPB	19 Jun 2020 15:48:05	U	SMPL
L3035-14 - 1	2-SP-3-5	-12	0.0307	-7470A	PPB	19 Jun 2020 15:50:12	U	SMPL
L3052-03 - 1	177	-12	0.0307	-7470A	PPB	19 Jun 2020 15:52:20	U	SMPL
PB129652BL - 1	PBW	-8	0.0317	-7470A	PPB	19 Jun 2020 15:54:28	U	SMPL
PB129652BS - 1	LCSW	15465	3.8017	-7470A	PPB	19 Jun 2020 15:56:34	U	SMPL
CCV54 - 1	CCV54	18920	4.6435	-7470A	PPB	19 Jun 2020 15:58:40	U	SMPL
CCB54 - 1	CCB54	-96	0.0103	-7470A	PPB	19 Jun 2020 16:00:47	U	SMPL
L3054-01 - 1	218	-64	0.0181	-7470A	PPB	19 Jun 2020 16:02:53	U	SMPL
L3054-01DUP - 1	218DUP	-84	0.0132	-7470A	PPB	19 Jun 2020 16:04:59	U	SMPL
L3054-01MS - 1	218MS	14890	3.6616	-7470A	PPB	19 Jun 2020 16:09:30	U	SMPL
L3054-01MSD - 1	218MSD	15090	3.7103	-7470A	PPB	19 Jun 2020 16:11:36	U	SMPL
PB129651TB - 1	PB129626TB	-93	0.0110	-7470A	PPB	19 Jun 2020 16:13:42	U	SMPL
L3034-03LX5 - 1		-21	0.0285	-7470A	PPB	19 Jun 2020 16:15:50	U	SMPL
L3034-03A - 1		15288	3.7586	-7470A	PPB	19 Jun 2020 16:17:57	U	SMPL
L3054-01LX5 - 1		-76	0.0151	-7470A	PPB	19 Jun 2020 16:20:04	U	SMPL
L3054-01A - 1		15625	3.8407	-7470A	PPB	19 Jun 2020 16:22:12	U	SMPL
CCV55 - 1	CCV55	19102	4.6879	-7470A	PPB	19 Jun 2020 16:24:19	U	SMPL
CCB55 - 1	CCB55	-73	0.0159	-7470A	PPB	19 Jun 2020 16:26:25	U	SMPL