



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109673

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/19/2020	17:00
2	CAL2	1	Water	NA	NA	20.2	7.01	06/19/2020	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	06/19/2020	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	06/19/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	06/19/2020	17:06
6	L3051-02	1	Solid	20.03	20	20.4	5.70	06/19/2020	17:09
7	L3081-01	1	Solid	20.04	20	20.5	8.57	06/19/2020	17:10
8	L3081-01DUP	1	Solid	20.03	20	22.4	8.58	06/19/2020	17:11
9	CCV2	1	Water	NA	NA	20.2	12.02	06/19/2020	17:15

WORKLIST(Hardcopy Internal Chain)

VB109673

WorkList Name : CORROSIVITY L3070 WorkList ID : 139780 Department : Wet-Chemistry Date : 06-19-2020 12:18:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2020	Solid	L3051-02	Corrosivity	Cool 4 deg C	PSEG01	P51	SLUDGE-BOX	06/18/2020	9045D
06/18/2020	Solid	L3081-01	Corrosivity	Cool 4 deg C	PARS15	P61	JFK-WC-S-SB-01	06/18/2020	9045D

Date/Time 06/19/2020 16:20
 Received by: JP
 Relinquished by: JP

Date/Time 06-19-2020 18:00
 Received by: JP
 Relinquished by: JP