



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109674

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/19/2020	17:00
2	CAL2	1	Water	NA	NA	20.2	7.01	06/19/2020	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	06/19/2020	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	06/19/2020	17:05
5	CCV1	1	Water	NA	NA	20.2	2.01	06/19/2020	17:10
6	L3081-02	1	Water	NA	NA	19.9	6.61	06/19/2020	17:25
7	L3081-02DUP	1	Water	NA	NA	19.8	6.62	06/19/2020	17:26
8	CCV2	1	Water	NA	NA	20.2	12.02	06/19/2020	17:30

WORKLIST(Hardcopy Internal Chain)

VB109674

WorkList Name : PH L3081 W WorkList ID : 139786 Department : Wet-Chemistry Date : 06-19-2020 16:32:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2020	Water	L3081-02	-D pH	Cool 4 deg C	PARS15	P61	JFK-WC-L-01	06/18/2020	9040C

Date/Time 06-19-2020 16:35
 Received by: Jo
 Relinquished by: Jo

Date/Time 06-19-2020 18:00
 Received by: AP
 Relinquished by: Jo