



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109683

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/22/2020	13:50
2	CAL2	1	Water	NA	NA	20.2	7.01	06/22/2020	13:51
3	CAL3	1	Water	NA	NA	20.2	10.03	06/22/2020	13:53
4	ICV	1	Water	NA	NA	20.2	7.00	06/22/2020	14:00
5	CCV1	1	Water	NA	NA	20.3	2.01	06/22/2020	14:02
6	L3083-01	1	Solid	20.03	20	22.9	5.68	06/22/2020	14:05
7	L3083-01DUP	1	Solid	20.04	20	22.8	5.69	06/22/2020	14:06
8	L3083-02	1	Solid	20.02	20	22.8	6.45	06/22/2020	14:09
9	L3083-03	1	Solid	20.03	20	22.6	8.67	06/22/2020	14:10
10	L3083-04	1	Solid	20.05	20	22.7	8.26	06/22/2020	14:11
11	CCV2	1	Water	NA	NA	20.3	12.02	06/22/2020	14:11

109683

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3083 WorkList ID : 139822 Department : Wet-Chemistry Date : 06-22-2020 13:46:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2020	Solid	L3083-01 ✓	Corrosivity	Cool 4 deg C	DIST01	P41	WC-1	06/19/2020	9045D
06/19/2020	Solid	L3083-02 ✓	Corrosivity	Cool 4 deg C	DIST01	P41	WC-2	06/19/2020	9045D
06/19/2020	Solid	L3083-03 ✓	Corrosivity	Cool 4 deg C	DIST01	P41	WC-3	06/19/2020	9045D
06/19/2020	Solid	L3083-04 ✓	Corrosivity	Cool 4 deg C	DIST01	P41	WC-4	06/19/2020	9045D

Date/Time 06/22/2020 13:40
 Received by: JP
 Relinquished by: JP

Date/Time 06/22/2020 15:00
 Received by: JP
 Relinquished by: JP