



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109695

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/23/2020	10:45
2	L3081-02		1	8	100.00	>212.0	>212.0	06/23/2020	11:10
3	L3081-02DUP		1	8	100.00	>212.0	>212.0	06/23/2020	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp62320 WorkList ID : 139844 Department : Wet-Chemistry Date : 06-23-2020 10:07:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2020	Water	L3081-02	☒ Flash Point	Cool 4 deg C	PARS15	P61	JFK-WC-L-01	06/18/2020	1010A

Date/Time 06/23/2020 10:15
 Received by: KS
 Relinquished by: AS

Date/Time 06/23/2020 11:45
 Received by: SP
 Relinquished by: KS