

LB 109701


7471B

INSTRUMENT ID: ~~OK~~ CX1NB
6/23/20

Linear

 μ Abs.:

55302



A= 0.0000e+000

B= 1.8217e-004

C= -8.9679e-002

Rho= 0.9998445

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.071	-0.071	103	0.000	103				
0.2	0.200	0.210	0.010	1644	0.0 %	1644				
2.5	2.500	2.530	0.030	14383	0.0 %	14383				
5.0	5.000	5.118	0.118	28587	0.0 %	28587				
7.5	7.500	7.428	-0.072	41267	0.0 %	41267				
10.0	10.000	9.985	-0.015	55302	0.0 %	55302				

LB109701 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Jun 2020 13:49:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	103	-	0.0000	7471B	PPB	23 Jun 2020 13:52:16	S	Std
0.2 - 1	50.2	1644	-	0.2000	7471B	PPB	23 Jun 2020 13:54:33	S	Std
2.5 - 1	52.5	14383	-	2.5000	7471B	PPB	23 Jun 2020 13:56:50	S	Std
5.0 - 1	55	28587	-	5.0000	7471B	PPB	23 Jun 2020 13:59:06	S	Std
7.5 - 1	57.5	41267	-	7.5000	7471B	PPB	23 Jun 2020 14:01:23	S	Std
10.0 - 1	510	55302	-	10.0000	7471B	PPB	23 Jun 2020 14:06:14	S	Std
ICV95 - 1	ICV95	23340	4.1622	-	7471B	PPB	23 Jun 2020 14:09:04	U	SMPL
ICB95 - 1	ICB95	-9	-0.0913	-	7471B	PPB	23 Jun 2020 14:11:19	U	SMPL
CCV65 - 1	CCV65	28817	5.1599	-	7471B	PPB	23 Jun 2020 14:13:34	U	SMPL
CCB65 - 1	CCB65	-11	-0.0917	-	7471B	PPB	23 Jun 2020 14:15:49	U	SMPL
CRA - 1	CRA	1665	0.2136	-	7471B	PPB	23 Jun 2020 14:18:05	U	SMPL
HighStd - 1	HighStd	55846	10.0838	-	7471B	PPB	23 Jun 2020 14:20:20	U	SMPL
ChkStd - 1	ChkStd	41884	7.5403	-	7471B	PPB	23 Jun 2020 14:22:36	U	SMPL
PB129705BL - 1	PBS	133	-0.0655	-	7471B	PPB	23 Jun 2020 14:24:53	U	SMPL
PB129705BS - 1	LCSS	21763	3.8749	-	7471B	PPB	23 Jun 2020 14:27:10	U	SMPL
L3082-01 - 1	SB-1	4096	0.6565	-	7471B	PPB	23 Jun 2020 14:34:23	U	SMPL
L3082-01DUP - 1	SB-1DUP	4058	0.6496	-	7471B	PPB	23 Jun 2020 14:36:39	U	SMPL
L3082-01MS - 1	SB-1MS	25494	4.5546	-	7471B	PPB	23 Jun 2020 14:38:56	U	SMPL
L3082-01MSD - 1	SB-1MSD	27303	4.8841	-	7471B	PPB	23 Jun 2020 14:41:13	U	SMPL
L3082-02 - 1	SB-2	3010	0.4587	-	7471B	PPB	23 Jun 2020 14:43:28	U	SMPL
CCV66 - 1	CCV66	28826	5.1616	-	7471B	PPB	23 Jun 2020 14:45:44	U	SMPL
CCB66 - 1	CCB66	0	-0.0897	-	7471B	PPB	23 Jun 2020 14:48:00	U	SMPL
L3082-03 - 1	SB-3	364	-0.0234	-	7471B	PPB	23 Jun 2020 14:50:15	U	SMPL
L3082-04 - 1	SB-4	533	0.0074	-	7471B	PPB	23 Jun 2020 14:52:31	U	SMPL
L3082-05 - 1	SB-5	2879	0.4348	-	7471B	PPB	23 Jun 2020 14:54:46	U	SMPL
L3082-06 - 1	SB-7	2153	0.3025	-	7471B	PPB	23 Jun 2020 14:57:01	U	SMPL
L3082-07 - 1	SB-8	1798	0.2379	-	7471B	PPB	23 Jun 2020 14:59:17	U	SMPL
L3082-08 - 1	SB-9	1139	0.1178	-	7471B	PPB	23 Jun 2020 15:01:34	U	SMPL
L3082-09 - 1	SB-10	6481	1.0910	-	7471B	PPB	23 Jun 2020 15:03:50	U	SMPL
L3082-10 - 1	SB-11	4289	0.6916	-	7471B	PPB	23 Jun 2020 15:06:07	U	SMPL
L3082-11 - 1	SB-12	116241	21.0859	-	7471B	PPB	23 Jun 2020 15:08:24	U	SMPL
L3082-12 - 1	SB-13	5961	0.9962	-	7471B	PPB	23 Jun 2020 15:10:40	U	SMPL
CCV67 - 1	CCV67	28710	5.1404	-	7471B	PPB	23 Jun 2020 15:12:58	U	SMPL
CCB67 - 1	CCB67	-14	-0.0922	-	7471B	PPB	23 Jun 2020 15:15:13	U	SMPL
L3082-13 - 1	SB-14	10659	1.8521	-	7471B	PPB	23 Jun 2020 15:17:29	U	SMPL
L3082-14 - 1	SB-15	10931	1.9016	-	7471B	PPB	23 Jun 2020 15:19:45	U	SMPL
L3082-15 - 1	SB-16	3586	0.5636	-	7471B	PPB	23 Jun 2020 15:22:00	U	SMPL
L3082-16 - 1	SB-17	25351	4.5285	-	7471B	PPB	23 Jun 2020 15:24:16	U	SMPL
L3082-17 - 1	SB-18	44688	8.0511	-	7471B	PPB	23 Jun 2020 15:26:32	U	SMPL
L3082-18 - 1	SB-19	1413	0.1677	-	7471B	PPB	23 Jun 2020 15:28:48	U	SMPL
L3082-19 - 1	SB-20	2837	0.4271	-	7471B	PPB	23 Jun 2020 15:31:05	U	SMPL
L3082-01LX5 - 1		781	0.0526	-	7471B	PPB	23 Jun 2020 15:33:22	U	SMPL
L3082-01A - 1		26084	4.6620	-	7471B	PPB	23 Jun 2020 15:35:39	U	SMPL
L3082-11DLX5 - 1		22875	4.0775	-	7471B	PPB	23 Jun 2020 15:37:55	U	SMPL
CCV68 - 1	CCV68	28635	5.1268	-	7471B	PPB	23 Jun 2020 15:40:10	U	SMPL
CCB68 - 1	CCB68	-3	-0.0902	-	7471B	PPB	23 Jun 2020 15:42:27	U	SMPL