



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109703

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/23/2020	14:10
2	CAL2	1	Water	NA	NA	20.2	7.01	06/23/2020	14:11
3	CAL3	1	Water	NA	NA	20.2	10.03	06/23/2020	14:12
4	ICV	1	Water	NA	NA	20.2	6.99	06/23/2020	14:15
5	CCV1	1	Water	NA	NA	20.2	2.02	06/23/2020	14:16
6	L3087-01	1	Water	NA	NA	23.1	8.50	06/23/2020	14:20
7	L3087-01DUP	1	Water	NA	NA	23.2	8.51	06/23/2020	14:21
8	CCV2	1	Water	NA	NA	20.1	12.02	06/23/2020	14:22

WORKLIST(Hardcopy Internal Chain)

18109703

WorkList Name : PH L3087

WorkList ID : 139849

Department : Wet-Chemistry

Date : 06-23-2020 13:58:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/23/2020	Water	L3087-01	pH	Cool 4 deg C	PSEG01	R11	FM1296	06/23/2020	9040C

Date/Time 06/23/2020 14:00
Received by: JPS
Relinquished by:

Date/Time 06/23/2020 15:00
Received by: KS
Relinquished by: JPS