



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109707

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/23/2020	17:35
2	CAL2	1	Water	NA	NA	20.2	7.00	06/23/2020	17:36
3	CAL3	1	Water	NA	NA	20.2	10.03	06/23/2020	17:37
4	ICV	1	Water	NA	NA	20.2	7.00	06/23/2020	17:39
5	CCV1	1	Water	NA	NA	20.2	2.01	06/23/2020	17:40
6	L2818-10	1	Solid	20.03	20	21.9	8.63	06/23/2020	17:42
7	L3089-02	1	Solid	20.04	20	24.6	7.50	06/23/2020	17:44
8	L3089-05	1	Solid	20.03	20	24.6	7.78	06/23/2020	17:45
9	L3089-08	1	Solid	20.02	20	24.7	7.80	06/23/2020	17:46
10	L3089-08DUP	1	Solid	20.03	20	24.8	7.81	06/23/2020	17:47
11	CCV2	1	Water	NA	NA	20.2	12.02	06/23/2020	17:49

VB 109707

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY L3086 WorkList ID: 139854 Department: Wet-Chemistry Date: 06-23-2020 14:57:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/23/2020	Solid	L2818-10	Corrosivity	Cool 4 deg C	PSEG05		TR-05-062320	06/23/2020	9045D
06/23/2020	Solid	L3089-02	Corrosivity	Cool 4 deg C	PSEG01	R12	RB47184-RT	06/23/2020	9045D
06/23/2020	Solid	L3089-05	Corrosivity	Cool 4 deg C	PSEG01	R12	RB44767-RT	06/23/2020	9045D
06/23/2020	Solid	L3089-08	Corrosivity	Cool 4 deg C	PSEG01	R12	RB47185-RT	06/23/2020	9045D

Date/Time 06-23-2020 17:20
Received by: JS
Relinquished by: AS

Date/Time 06-23-2020 18:35
Received by: JS
Relinquished by: JS