



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109712

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/24/2020	08:35
2	CAL2	1	Water	NA	NA	20.1	7.01	06/24/2020	08:36
3	CAL3	1	Water	NA	NA	20.1	10.03	06/24/2020	08:37
4	ICV	1	Water	NA	NA	20.1	6.98	06/24/2020	08:40
5	CCV1	1	Water	NA	NA	20.1	2.01	06/24/2020	08:42
6	L3085-01	1	Solid	20.03	20	22.1	7.17	06/24/2020	08:45
7	L3085-02	1	Solid	20.02	20	22.5	6.75	06/24/2020	08:46
8	L3085-02DUP	1	Solid	20.03	20	22.6	6.76	06/24/2020	08:47
9	CCV2	1	Water	NA	NA	20.2	12.02	06/24/2020	08:50

13109712

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH L3085 WorkList ID: 139867 Department: Wet-Chemistry Date: 06-24-2020 08:29:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/23/2020	Solid	L3085-01	pH	Cool 4 deg C	PSEG01	R12	31063	06/23/2020	9045D
06/23/2020	Solid	L3085-02	pH	Cool 4 deg C	PSEG01	R12	31064	06/23/2020	9045D

Date/Time 06/24/2020 8:30
Received by: Jn
Relinquished by: Jn

Date/Time 06/24/2020 9:45
Received by: Jn
Relinquished by: Jn