



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109721

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.00      | 80.6      | 80.6            | 06/24/2020 | 09:35     |
| 2   | L3087-02    |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 06/24/2020 | 10:10     |
| 3   | L3087-02DUP |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 06/24/2020 | 10:40     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB109721

WorkList Name : fp62420      WorkList ID : 139875      Department : Wet-Chemistry      Date : 06-24-2020 09:02:05

| Due Date   | Matrix | Sample   | Test          | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|---------------|--------------|----------|------------------|-----------------|--------------|--------|
| 06/26/2020 | Water  | L3087-02 | C Flash Point | Cool 4 deg C | PSEG01   | R11              | FM1296          | 06/23/2020   | 1010A  |

Date/Time 06/24/20 09:05  
 Received by: KS  
 Relinquished by: JP

Date/Time 06/24/20 10:45  
 Received by: JP  
 Relinquished by: KS