



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109728

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/25/2020	10:10
2	L3098-01		1	8	100.00	>212.0	>212.0	06/25/2020	10:45
3	L3098-01DUP		1	8	100.00	>212.0	>212.0	06/25/2020	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB109728

WorkList Name : fp62520 WorkList ID : 139902 Department : Wet-Chemistry Date : 06-25-2020 09:34:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2020	Water	L3098-01	F	Flash Point	Cool 4 deg C	PSEG01	R21	31106-31128	06/24/2020 1010A

Date/Time 06/25/20 09:40
 Received by: KS
 Relinquished by: JS

Date/Time 06/25/20 11:25
 Received by: JP
 Relinquished by: KS