

LB109729


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5612E



A= 0.0000e+000

B= 1.8005e-004

C= 1.7551e-002

Rho= 0.9997339

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	1	0.000	1				
0.2	0.200	0.218	0.018	1111	0.0 %	1111				
2.5	2.500	2.535	0.035	13984	0.0 %	13984				
5.0	5.000	4.965	-0.035	27479	0.0 %	27479				
7.5	7.500	7.341	-0.159	40675	0.0 %	40675				
10.0	10.000	10.123	0.123	56126	0.0 %	56126				

LB109729 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 25 Jun 2020 14:03:02					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.0000	7471B	PPB	25 Jun 2020 14:04:14	S	Std
0.2 - 1	50.2	1111	-	0.2000	7471B	PPB	25 Jun 2020 14:06:31	S	Std
2.5 - 1	52.5	13984	-	2.5000	7471B	PPB	25 Jun 2020 14:08:48	S	Std
5.0 - 1	55	27479	-	5.0000	7471B	PPB	25 Jun 2020 14:11:05	S	Std
7.5 - 1	57.5	40675	-	7.5000	7471B	PPB	25 Jun 2020 14:13:21	S	Std
10.0 - 1	510	56126	-	10.0000	7471B	PPB	25 Jun 2020 14:15:37	S	Std
ICV99 - 1	ICV99	21425	3.8751	-7471B	PPB	25 Jun 2020 14:18:29	U	SMPL	
ICB99 - 1	ICB99	-106	-0.0015	-7471B	PPB	25 Jun 2020 14:20:44	U	SMPL	
CCV81 - 1	CCV81	27463	4.9623	-7471B	PPB	25 Jun 2020 14:22:59	U	SMPL	
CCB81 - 1	CCB81	-114	-0.0030	-7471B	PPB	25 Jun 2020 14:25:14	U	SMPL	
CRA - 1	CRA	1063	0.2089	-7471B	PPB	25 Jun 2020 14:27:30	U	SMPL	
HighStd - 1	HighStd	56264	10.1479	-7471B	PPB	25 Jun 2020 14:29:45	U	SMPL	
ChkStd - 1	ChkStd	37338	6.7403	-7471B	PPB	25 Jun 2020 14:32:00	U	SMPL	
PB129755BL - 1	PBS	-114	-0.0030	-7471B	PPB	25 Jun 2020 14:34:18	U	SMPL	
PB129755BS - 1	LCSS	21663	3.9180	-7471B	PPB	25 Jun 2020 14:36:35	U	SMPL	
L2810-09 - 1	CL-01-062420	7359	1.3425	-7471B	PPB	25 Jun 2020 14:43:47	U	SMPL	
L2810-09DUP - 1	CL-01-062420DUP	7327	1.3368	-7471B	PPB	25 Jun 2020 14:46:03	U	SMPL	
L2810-09MS - 1	CL-01-062420MS	30936	5.5876	-7471B	PPB	25 Jun 2020 14:48:20	U	SMPL	
L2810-09MSD - 1	CL-01-062420MSD	28956	5.2311	-7471B	PPB	25 Jun 2020 14:50:37	U	SMPL	
L2815-11 - 1	OR-02-062420	3019	0.5611	-7471B	PPB	25 Jun 2020 14:52:52	U	SMPL	
CCV82 - 1	CCV82	27240	4.9221	-7471B	PPB	25 Jun 2020 14:55:08	U	SMPL	
CCB82 - 1	CCB82	-92	0.0010	-7471B	PPB	25 Jun 2020 14:57:23	U	SMPL	
L2817-11 - 1	SU-01-062420	2372	0.4446	-7471B	PPB	25 Jun 2020 14:59:39	U	SMPL	
L3099-01 - 1	DUCT-BANK	1713	0.3260	-7471B	PPB	25 Jun 2020 15:01:54	U	SMPL	
L3099-03 - 1	TP-4	1213	0.2360	-7471B	PPB	25 Jun 2020 15:04:10	U	SMPL	
L3100-01 - 1	72-11934-VNJ-230	853	0.1711	-7471B	PPB	25 Jun 2020 15:06:25	U	SMPL	
L3101-01 - 1	EBS8-S-01	5109	0.9374	-7471B	PPB	25 Jun 2020 15:08:41	U	SMPL	
L3101-02 - 1	EBS8-S-02	5147	0.9443	-7471B	PPB	25 Jun 2020 15:10:58	U	SMPL	
L3101-03 - 1	EBS8-S-02A	4520	0.8314	-7471B	PPB	25 Jun 2020 15:13:14	U	SMPL	
L3101-04 - 1	EBS8-S-03	5162	0.9470	-7471B	PPB	25 Jun 2020 15:15:31	U	SMPL	
L3111-03 - 1	TP-6-S	1514	0.2901	-7471B	PPB	25 Jun 2020 15:17:48	U	SMPL	
L3111-08 - 1	TP-6-D	442	0.0971	-7471B	PPB	25 Jun 2020 15:20:03	U	SMPL	
CCV83 - 1	CCV83	27572	4.9819	-7471B	PPB	25 Jun 2020 15:22:19	U	SMPL	
CCB83 - 1	CCB83	-101	-0.0006	-7471B	PPB	25 Jun 2020 15:24:34	U	SMPL	
L2810-09LX5 - 1		1455	0.2795	-7471B	PPB	25 Jun 2020 15:26:50	U	SMPL	
L2810-09A - 1		30225	5.4596	-7471B	PPB	25 Jun 2020 15:29:06	U	SMPL	
CCV84 - 1	CCV84	27543	4.9767	-7471B	PPB	25 Jun 2020 15:31:21	U	SMPL	
CCB84 - 1	CCB84	-95	0.0004	-7471B	PPB	25 Jun 2020 15:33:38	U	SMPL	