


LB109745

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	-46	0.000	-46				
0.2	0.200	0.219	0.019	708	0.0 %	708				
2.5	2.500	2.446	-0.054	9011	0.0 %	9011				
5.0	5.000	4.942	-0.058	18320	0.0 %	18320				
7.5	7.500	7.651	0.151	28425	0.0 %	28425				
10.0	10.000	9.925	-0.075	36904	0.0 %	36904				

LB109745 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jun 2020 08:56:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-46	-	0.0000	7470A	PPB	26 Jun 2020 09:10:39	S	Std
0.2 - 1	50.2	708	-	0.2000	7470A	PPB	26 Jun 2020 09:12:47	S	Std
2.5 - 1	52.5	9011	-	2.5000	7470A	PPB	26 Jun 2020 09:14:55	S	Std
5.0 - 1	55	18320	-	5.0000	7470A	PPB	26 Jun 2020 09:17:02	S	Std
7.5 - 1	57.5	28425	-	7.5000	7470A	PPB	26 Jun 2020 09:19:10	S	Std
10.0 - 1	510	36904	-	10.0000	7470A	PPB	26 Jun 2020 09:24:04	S	Std
ICV01 - 1	ICV01	15383	4.1542	-	7470A	PPB	26 Jun 2020 09:26:40	U	SMPL
ICB01 - 1	ICB01	-115	-0.0016	-	7470A	PPB	26 Jun 2020 09:28:46	U	SMPL
CCV85 - 1	CCV85	18284	4.9321	-	7470A	PPB	26 Jun 2020 09:30:52	U	SMPL
CCB85 - 1	CCB85	-112	-0.0008	-	7470A	PPB	26 Jun 2020 09:32:57	U	SMPL
CRA - 1	CRA	696	0.2159	-	7470A	PPB	26 Jun 2020 09:35:03	U	SMPL
HighStd - 1	HighStd	36702	9.8710	-	7470A	PPB	26 Jun 2020 09:37:09	U	SMPL
ChkStd - 1	ChkStd	26181	7.0498	-	7470A	PPB	26 Jun 2020 09:39:15	U	SMPL
PB129777BL - 1	PBW	-170	-0.0163	-	7470A	PPB	26 Jun 2020 09:41:23	U	SMPL
PB129777BS - 1	LCSW	15154	4.0928	-	7470A	PPB	26 Jun 2020 09:43:30	U	SMPL
L2810-16 - 1	CL-01-062420	-97	0.0032	-	7470A	PPB	26 Jun 2020 09:45:37	U	SMPL
L2810-16DUP - 1	CL-01-062420DUP	-44	0.0174	-	7470A	PPB	26 Jun 2020 09:47:44	U	SMPL
L2810-16MS - 1	CL-01-062420MS	14992	4.0494	-	7470A	PPB	26 Jun 2020 09:49:51	U	SMPL
L2810-16MSD - 1	CL-01-062420MSD	15186	4.1014	-	7470A	PPB	26 Jun 2020 09:51:59	U	SMPL
L2815-15 - 1	OR-02-062420	-96	0.0035	-	7470A	PPB	26 Jun 2020 09:54:05	U	SMPL
CCV86 - 1	CCV86	18814	5.0743	-	7470A	PPB	26 Jun 2020 09:56:11	U	SMPL
CCB86 - 1	CCB86	-114	-0.0013	-	7470A	PPB	26 Jun 2020 09:58:17	U	SMPL
L2817-14 - 1	SU-01-062420	-42	0.0180	-	7470A	PPB	26 Jun 2020 10:00:23	U	SMPL
L3111-04 - 1	TP-6-S	-25	0.0225	-	7470A	PPB	26 Jun 2020 10:02:30	U	SMPL
L3111-08 - 1	TP-6-D	-30	0.0212	-	7470A	PPB	26 Jun 2020 10:04:36	U	SMPL
PB129778BL - 1	PBW	-19	0.0241	-	7470A	PPB	26 Jun 2020 10:06:42	U	SMPL
PB129778BS - 1	LCSW	13814	3.7335	-	7470A	PPB	26 Jun 2020 10:08:50	U	SMPL
L3098-01 - 1	31106-31128	-76	0.0089	-	7470A	PPB	26 Jun 2020 10:10:57	U	SMPL
L3098-01DUP - 1	31106-31128DUP	-36	0.0196	-	7470A	PPB	26 Jun 2020 10:13:05	U	SMPL
L3098-01MS - 1	31106-31128MS	15588	4.2092	-	7470A	PPB	26 Jun 2020 10:15:12	U	SMPL
L3098-01MSD - 1	31106-31128MSD	15414	4.1625	-	7470A	PPB	26 Jun 2020 10:17:20	U	SMPL
CCV87 - 1	CCV87	18871	5.0896	-	7470A	PPB	26 Jun 2020 10:19:26	U	SMPL
CCB87 - 1	CCB87	-125	-0.0043	-	7470A	PPB	26 Jun 2020 10:21:33	U	SMPL
PB129777TB - 1	PB129748TB	-48	0.0164	-	7470A	PPB	26 Jun 2020 10:23:39	U	SMPL
PB129778TB - 1	PB129765TB	-41	0.0182	-	7470A	PPB	26 Jun 2020 10:25:45	U	SMPL
L2810-16LX5 - 1		-41	0.0182	-	7470A	PPB	26 Jun 2020 10:27:51	U	SMPL
L2810-16A - 1		15402	4.1593	-	7470A	PPB	26 Jun 2020 10:29:58	U	SMPL
L3098-01LX5 - 1		-105	0.0011	-	7470A	PPB	26 Jun 2020 10:32:04	U	SMPL
L3098-01A - 1		15479	4.1800	-	7470A	PPB	26 Jun 2020 10:34:11	U	SMPL
CCV88 - 1	CCV88	18620	5.0222	-	7470A	PPB	26 Jun 2020 10:36:18	U	SMPL
CCB88 - 1	CCB88	-99	0.0027	-	7470A	PPB	26 Jun 2020 10:38:25	U	SMPL