



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109746

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	06/25/2020	17:30
2	CAL2	1	Water	NA	NA	21.3	7.00	06/25/2020	17:31
3	CAL3	1	Water	NA	NA	21.4	10.03	06/25/2020	17:32
4	ICV	1	Water	NA	NA	21.2	6.99	06/25/2020	17:35
5	CCV1	1	Water	NA	NA	21.2	2.02	06/25/2020	17:36
6	L3111-10	1	Solid	20.03	20	21.7	7.05	06/25/2020	17:38
7	L3113-02	1	Solid	20.03	20	21.9	6.40	06/25/2020	17:40
8	L3114-02	1	Solid	20.01	20	21.8	7.33	06/25/2020	17:42
9	L3115-02	1	Solid	20.03	20	23.8	5.64	06/25/2020	17:44
10	L3126-04	1	Solid	20.02	20	21.3	8.34	06/25/2020	17:45
11	L3126-08	1	Solid	20.03	20	21.8	8.52	06/25/2020	17:47
12	L3126-12	1	Solid	20.03	20	21.2	6.08	06/25/2020	17:48
13	L3126-12DUP	1	Solid	20.04	20	21.3	6.09	06/25/2020	17:50
14	CCV2	1	Water	NA	NA	21.3	12.02	06/25/2020	18:00

✓ 109746

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COROSIVITY L3113

WorkList ID : 139922

Department : Wet-Chemistry

Date : 06-25-2020 13:21:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2020	Solid	L3111-10	Corrosivity	Cool 4 deg C	PSEG01	R22	TP-6-D	06/24/2020	9045D
06/25/2020	Solid	L3113-02	Corrosivity	Cool 4 deg C	PSEG01	R22	DRUM-COMP	06/25/2020	9045D
06/25/2020	Solid	L3114-02	Corrosivity	Cool 4 deg C	PSEG01	R22	VNJ-242	06/25/2020	9045D
06/25/2020	Solid	L3115-02	Corrosivity	Cool 4 deg C	PSEG01	R22	TP	06/25/2020	9045D
06/25/2020	Solid	L3126-04	Corrosivity	Cool 4 deg C	PSEG01	R22	TP-8-S	06/25/2020	9045D
06/25/2020	Solid	L3126-08	Corrosivity	Cool 4 deg C	PSEG01	R22	TP-7-S	06/25/2020	9045D
06/25/2020	Solid	L3126-12	Corrosivity	Cool 4 deg C	PSEG01	R22	TP-4-S	06/25/2020	9045D

Date/Time 06/25/2020 17:00
 Received by: JP
 Relinquished by: KS

Date/Time 06/25/2020 18:20
 Received by: KS
 Relinquished by: JP