



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109747

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.00      | 80.6      | 80.6            | 06/26/2020 | 10:35     |
| 2   | L3125-01    |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 06/26/2020 | 11:05     |
| 3   | L3125-01DUP |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 06/26/2020 | 11:40     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB 109747

WorkList Name : fp62620      WorkList ID : 139949      Department : Wet-Chemistry      Date : 06-26-2020 09:58:11

| Due Date   | Matrix | Sample   | Test          | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|---------------|--------------|----------|------------------|-----------------|--------------|--------|
| 07/02/2020 | Water  | L3125-01 | K Flash Point | Cool 4 deg C | ENTA05   | R41              | TW-2            | 06/24/2020   | 1010A  |

Date/Time 06/26/2020 10:00  
 Received by: KS (wet chem)  
 Relinquished by: JQ (wet chem)

Date/Time 06/26/2020 11:45  
 Received by: JD (wet chem)  
 Relinquished by: KS (wet chem)