



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109753

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/26/2020	11:48
2	CAL2	1	Water	NA	NA	20.1	7.00	06/26/2020	11:49
3	CAL3	1	Water	NA	NA	20.1	10.03	06/26/2020	11:50
4	ICV	1	Water	NA	NA	21.2	7.02	06/26/2020	11:52
5	CCV1	1	Water	NA	NA	21.1	2.02	06/26/2020	12:00
6	L3138-01	1	Water	NA	NA	22.2	4.55	06/26/2020	12:02
7	L3138-01DUP	1	Water	NA	NA	22.3	4.56	06/26/2020	12:03
8	CCV2	1	Water	NA	NA	21.1	12.02	06/26/2020	12:05

103109753

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3138 WorkList ID : 139955 Department : Wet-Chemistry Date : 06-26-2020 11:44:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/25/2020	Water	L3138-01-17	pH	Cool 4 deg C	STER13	R31	CELL-B-POOL-WATER	06/25/2020	SM 4500-H B

Date/Time 06/26/2020 11:45
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 06/26/2020 12:13
Received by: [Signature]
Relinquished by: [Signature]