



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109767

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/26/2020	15:35
2	CAL2	1	Water	NA	NA	20.2	6.99	06/26/2020	15:36
3	CAL3	1	Water	NA	NA	20.2	10.03	06/26/2020	15:37
4	ICV	1	Water	NA	NA	20.2	7.00	06/26/2020	15:39
5	CCV1	1	Water	NA	NA	20.1	2.02	06/26/2020	15:40
6	L3143-01	1	Water	NA	NA	11.8	6.87	06/26/2020	15:42
7	L3143-01DUP	1	Water	NA	NA	11.9	6.88	06/26/2020	15:43
8	CCV2	1	Water	NA	NA	20.2	12.02	06/26/2020	15:45

13109767

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3143 W WorkList ID : 139965 Department : Wet-Chemistry Date : 06-26-2020 15:25:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2020	Water	L3143-01 - F	Corrosivity	Cool 4 deg C	PARS02	R21	WC-20200626	06/26/2020	9040C

Date/Time 06/26/2020 15:30
Received by: Jz (wet-chem)
Relinquished by: Jz (wet-chem)

Date/Time 06/26/2020 16:20
Received by: Jz (wet-chem)
Relinquished by: Jz (wet-chem)