



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109769

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/26/2020	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	06/26/2020	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	06/26/2020	16:18
4	ICV	1	Water	NA	NA	20.3	6.99	06/26/2020	16:20
5	CCV1	1	Water	NA	NA	20.3	2.01	06/26/2020	16:22
6	L3145-01	1	Solid	20.03	20	21.3	7.60	06/26/2020	16:23
7	L3145-01DUP	1	Solid	20.04	20	21.4	7.61	06/26/2020	16:24
8	CCV2	1	Water	NA	NA	20.3	12.02	06/26/2020	16:25

VB109769

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3145 WorkList ID : 139966 Department : Wet-Chemistry Date : 06-26-2020 16:01:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2020	Solid	L3145-01	pH	Cool 4 deg C	WEST04	R31	MC0AA2	06/24/2020	9045D

Date/Time 06/24/2020 11:10
Received by: JG (wet chem)
Relinquished by: JG (wet chem)

Date/Time 06/24/2020 17:20
Received by: JG (wet chem)
Relinquished by: JG (wet chem)