



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109770

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	06/26/2020	18:00
2	CAL2	1	Water	NA	NA	20.1	7.00	06/26/2020	18:01
3	CAL3	1	Water	NA	NA	20.1	10.03	06/26/2020	18:02
4	ICV	1	Water	NA	NA	20.3	7.00	06/26/2020	18:05
5	CCV1	1	Water	NA	NA	20.1	2.01	06/26/2020	18:03
6	L3129-01	1	Water	NA	NA	26.0	7.52	06/26/2020	18:05
7	L3129-02	1	Water	NA	NA	25.1	6.90	06/26/2020	18:06
8	L3129-04	1	Water	NA	NA	25.4	6.55	06/26/2020	18:10
9	L3129-05	1	Water	NA	NA	25.6	6.20	06/26/2020	18:13
10	L3129-11	1	Water	NA	NA	24.3	6.02	06/26/2020	18:15
11	L3129-11DUP	1	Water	NA	NA	24.4	6.03	06/26/2020	18:16
12	CCV2	1	Water	NA	NA	20.1	12.02	06/26/2020	18:19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph L3129 WorkList ID : 139967 Department : Wet-Chemistry Date : 06-26-2020 16:34:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2020	Water	L3129-01	pH	Cool 4 deg C	PSEG01	R31	1446	06/26/2020	9040C
06/26/2020	Water	L3129-02	pH	Cool 4 deg C	PSEG01	R31	1445	06/26/2020	9040C
06/26/2020	Water	L3129-04	pH	Cool 4 deg C	PSEG01	R31	1405	06/26/2020	9040C
06/26/2020	Water	L3129-05	pH	Cool 4 deg C	PSEG01	R31	1406	06/26/2020	9040C
06/26/2020	Water	L3129-11	pH	Cool 4 deg C	PSEG01	R31	1308	06/26/2020	9040C

Date/Time 06/26/2020 16:35
 Received by: JS
 Relinquished by: JS

Date/Time 06/26/2020 18:00
 Received by: KS
 Relinquished by: JS