



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109772

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/26/2020	18:10
2	CAL2	1	Water	NA	NA	20.3	7.00	06/26/2020	18:11
3	CAL3	1	Water	NA	NA	20.3	10.03	06/26/2020	18:12
4	ICV	1	Water	NA	NA	20.4	7.02	06/26/2020	18:15
5	CCV1	1	Water	NA	NA	20.1	2.02	06/26/2020	18:20
6	L2806-04	1	Solid	20.03	20	23.1	9.72	06/26/2020	18:22
7	L2819-08	1	Solid	20.04	20	22.5	8.46	06/26/2020	18:25
8	L2819-08DUP	1	Solid	20.03	20	22.6	8.47	06/26/2020	18:26
9	CCV2	1	Water	NA	NA	20.3	12.02	06/26/2020	18:29

109772

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2806 WorkList ID : 139964 Department : Wet-Chemistry Date : 06-26-2020 14:49:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2020	Solid	L2806-04	Corrosivity	Cool 4 deg C	PSEG05	R31	AU-06-062620	06/26/2020	9045D
06/26/2020	Solid	L2819-08	Corrosivity	Cool 4 deg C	PSEG05		BU-03-062620	06/26/2020	9045D

Date/Time 06/26/2020 17:20
Received by: [Signature]
Relinquished by: KS (Hxm)

Date/Time 06/26/2020 19:00
Received by: KS
Relinquished by: [Signature]