


LB109773

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.041	-0.041	-50	0.000	-50				
0.2	0.200	0.171	-0.029	788	0.0 %	788				
2.5	2.500	2.557	0.057	10218	0.0 %	10218				
5.0	5.000	5.002	0.002	19880	0.0 %	19880				
7.5	7.500	7.617	0.117	30212	0.0 %	30212				
10.0	10.000	9.894	-0.106	39209	0.0 %	39209				

LB109773 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Jun 2020 09:28:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-50	-	0.0000	7470A	PPB	29 Jun 2020 10:56:59	S	Std
0.2 - 1	50.2	788	-	0.2000	7470A	PPB	29 Jun 2020 10:59:07	S	Std
2.5 - 1	52.5	10218	-	2.5000	7470A	PPB	29 Jun 2020 11:01:15	S	Std
5.0 - 1	55	19880	-	5.0000	7470A	PPB	29 Jun 2020 11:03:23	S	Std
7.5 - 1	57.5	30212	-	7.5000	7470A	PPB	29 Jun 2020 11:05:30	S	Std
10.0 - 1	57.11	39209	-	10.0000	7470A	PPB	29 Jun 2020 11:10:12	S	Std
ICV03 - 1	ICV03	16487	4.1437	-	7470A	PPB	29 Jun 2020 11:12:59	U	SMPL
ICB03 - 1	ICB03	-150	-0.0668	-	7470A	PPB	29 Jun 2020 11:15:04	U	SMPL
CCV92 - 1	CCV92	20078	5.0525	-	7470A	PPB	29 Jun 2020 11:17:10	U	SMPL
CCB92 - 1	CCB92	-155	-0.0680	-	7470A	PPB	29 Jun 2020 11:19:16	U	SMPL
CRA - 1	CRA	790	0.1711	-	7470A	PPB	29 Jun 2020 11:21:22	U	SMPL
HighStd - 1	HighStd	39133	9.8749	-	7470A	PPB	29 Jun 2020 11:23:28	U	SMPL
ChkStd - 1	ChkStd	28358	7.1480	-	7470A	PPB	29 Jun 2020 11:25:34	U	SMPL
PB129813BL - 1	PBW	-163	-0.0701	-	7470A	PPB	29 Jun 2020 11:27:42	U	SMPL
PB129813BS - 1	LCSW	15143	3.8036	-	7470A	PPB	29 Jun 2020 11:29:50	U	SMPL
L3113-02 - 1	DRUM-COMP	-107	-0.0559	-	7470A	PPB	29 Jun 2020 11:31:57	U	SMPL
L3113-02DUP - 1	DRUM-COMPDUP	-68	-0.0460	-	7470A	PPB	29 Jun 2020 11:34:05	U	SMPL
L3113-02MS - 1	DRUM-COMPMS	14811	3.7195	-	7470A	PPB	29 Jun 2020 11:36:12	U	SMPL
L3113-02MSD - 1	DRUM-COMPMSD	15415	3.8724	-	7470A	PPB	29 Jun 2020 11:38:20	U	SMPL
L3114-02 - 1	VNJ-242	-122	-0.0597	-	7470A	PPB	29 Jun 2020 11:40:25	U	SMPL
CCV93 - 1	CCV93	20650	5.1973	-	7470A	PPB	29 Jun 2020 11:42:31	U	SMPL
CCB93 - 1	CCB93	-132	-0.0622	-	7470A	PPB	29 Jun 2020 11:44:38	U	SMPL
L3115-02 - 1	TP	-66	-0.0455	-	7470A	PPB	29 Jun 2020 11:46:44	U	SMPL
L3126-04 - 1	TP-8-S	-41	-0.0392	-	7470A	PPB	29 Jun 2020 11:48:50	U	SMPL
L3126-08 - 1	TP-7-S	-46	-0.0405	-	7470A	PPB	29 Jun 2020 11:50:57	U	SMPL
L3126-12 - 1	TP-4-S	-60	-0.0440	-	7470A	PPB	29 Jun 2020 11:53:03	U	SMPL
PB129813TB - 1	PB129780TB	-64	-0.0450	-	7470A	PPB	29 Jun 2020 11:55:10	U	SMPL
L3113-02LX5 - 1		-57	-0.0432	-	7470A	PPB	29 Jun 2020 11:57:17	U	SMPL
L3113-02A - 1		15707	3.9463	-	7470A	PPB	29 Jun 2020 11:59:24	U	SMPL
CCV94 - 1	CCV94	20408	5.1360	-	7470A	PPB	29 Jun 2020 12:01:32	U	SMPL
CCB94 - 1	CCB94	-135	-0.0630	-	7470A	PPB	29 Jun 2020 12:03:40	U	SMPL