



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109776

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/25/2020	16:20
2	CAL2	1	Water	NA	NA	20.1	6.99	06/25/2020	16:21
3	CAL3	1	Water	NA	NA	20.1	10.03	06/25/2020	16:22
4	ICV	1	Water	NA	NA	20.1	7.00	06/25/2020	16:23
5	CCV1	1	Water	NA	NA	20.2	2.01	06/25/2020	16:25
6	L3125-01	1	Water	NA	NA	22.2	11.40	06/25/2020	16:26
7	L3125-01DUP	1	Water	NA	NA	22.3	11.41	06/25/2020	16:27
8	CCV2	1	Water	NA	NA	20.3	12.02	06/25/2020	16:30

2B 109776

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3125 WorkList ID : 140002 Department : Wet-Chemistry Date : 06-25-2020 16:10:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/24/2020	Water	L3125-01	pH	Cool 4 deg C	ENTA05	R41	TW-2	06/24/2020	9040C

Date/Time 06/25/2020 16:15
Received by: JP (wdr-hdm)
Relinquished by: AM (wetchen)

Date/Time 06/25/2020 17:00
Received by: AM (wetchen)
Relinquished by: JP (wdr-hdm)