

LB 109781


7471B

INSTRUMENT ID: CY 2

Linear

 μ Abs.:

57009



A= 0.0000e+000

B= 1.7523e-004

C= -7.8682e-003

Rho= 0.9999831

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-24	0.000	-24				
0.2	0.200	0.196	-0.004	1166	0.0 %	1166				
2.5	2.500	2.527	0.027	14465	0.0 %	14465				
5.0	5.000	4.976	-0.024	28441	0.0 %	28441				
7.5	7.500	7.531	0.031	43024	0.0 %	43024				
10.0	10.000	9.982	-0.018	57009	0.0 %	57009				

LB109781 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Jun 2020 13:40:16

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.00007471B	PPB	29 Jun 2020 13:42:17	S	Std
0.2 - 1	50.2	1166	-	0.20007471B	PPB	29 Jun 2020 13:44:34	S	Std
2.5 - 1	52.5	14465	-	2.50007471B	PPB	29 Jun 2020 13:46:51	S	Std
5.0 - 1	55	28441	-	5.00007471B	PPB	29 Jun 2020 13:49:08	S	Std
7.5 - 1	57.5	43024	-	7.50007471B	PPB	29 Jun 2020 13:51:24	S	Std
10.0 - 1	59	57009	-	10.00007471B	PPB	29 Jun 2020 13:53:41	S	Std
ICV04 - 1	ICV04	23078	4.0361	-7471B	PPB	29 Jun 2020 13:56:38	U	SMPL
ICB04 - 1	ICB04	-166	-0.0370	-7471B	PPB	29 Jun 2020 13:58:53	U	SMPL
CCV95 - 1	CCV95	28389	4.9667	-7471B	PPB	29 Jun 2020 14:01:07	U	SMPL
CCB95 - 1	CCB95	-167	-0.0371	-7471B	PPB	29 Jun 2020 14:03:22	U	SMPL
CRA - 1	CRA	1119	0.1882	-7471B	PPB	29 Jun 2020 14:05:38	U	SMPL
HighStd - 1	HighStd	56306	9.8586	-7471B	PPB	29 Jun 2020 14:07:53	U	SMPL
ChkStd - 1	ChkStd	40613	7.1087	-7471B	PPB	29 Jun 2020 14:10:09	U	SMPL
PB129838BL - 1	PBS	-189	-0.0410	-7471B	PPB	29 Jun 2020 14:12:26	U	SMPL
PB129838BS - 1	LCSS	20218	3.5349	-7471B	PPB	29 Jun 2020 14:14:43	U	SMPL
L2806-01 - 1	AU-06-062620	1289	0.2180	-7471B	PPB	29 Jun 2020 14:16:59	U	SMPL
L2819-06 - 1	BU-03-062620	3927	0.6803	-7471B	PPB	29 Jun 2020 14:24:53	U	SMPL
L2819-06DUP - 1	BU-03-062620DUP	4240	0.7351	-7471B	PPB	29 Jun 2020 14:29:42	U	SMPL
L2819-06MS - 1	BU-03-062620MS	24350	4.2590	-7471B	PPB	29 Jun 2020 14:31:59	U	SMPL
L2819-06MSD - 1	BU-03-062620MSD	24796	4.3371	-7471B	PPB	29 Jun 2020 14:34:14	U	SMPL
CCV96 - 1	CCV96	28579	5.0000	-7471B	PPB	29 Jun 2020 14:36:29	U	SMPL
CCB96 - 1	CCB96	-176	-0.0387	-7471B	PPB	29 Jun 2020 14:38:45	U	SMPL
L2819-06LX5 - 1		798	0.1320	-7471B	PPB	29 Jun 2020 14:41:01	U	SMPL
L2819-06A - 1		25962	4.5414	-7471B	PPB	29 Jun 2020 14:43:16	U	SMPL
CCV97 - 1	CCV97	28302	4.9515	-7471B	PPB	29 Jun 2020 14:45:32	U	SMPL
CCB97 - 1	CCB97	-187	-0.0406	-7471B	PPB	29 Jun 2020 14:47:47	U	SMPL