



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109782

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/29/2020	09:45
2	L3143-01		1	8	100.00	>212.0	>212.0	06/29/2020	10:10
3	L3143-01DUP		1	8	100.00	>212.0	>212.0	06/29/2020	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB109782

WorkList Name : fp62920 WorkList ID : 139994 Department : Wet-Chemistry Date : 06-29-2020 09:05:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2020	Water	L3143-01	Flash Point	Cool 4 deg C	PARS02	R21	WC-20200626	06/26/2020	1010A

Date/Time 06/29/2020 09:15
 Received by: KS (wetchem) *(wetchem)*
 Relinquished by: JP (wet chem)

Date/Time 06/29/2020 10:45
 Received by: JP (wetchem)
 Relinquished by: KS *(wetchem)*