

LB109783


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37892



A= 0.0000e+000

B= 2.6541e-004

C= 4.5376e-003

Rho= 0.9998335

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	-6	0.000	-6				
0.05	0.050	0.052	0.002	179	0.0 %	179				
0.2	0.200	0.186	-0.014	683	0.0 %	683				
2.5	2.500	2.508	0.008	9434	0.0 %	9434				
5.0	5.000	5.084	0.084	19138	0.0 %	19138				
7.5	7.500	7.355	-0.145	27695	0.0 %	27695				
10.0	10.000	10.062	0.062	37892	0.0 %	37892				

LB109783 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Jun 2020 15:17:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.0000	245.1	PPB	29 Jun 2020 15:18:58	S	Std
0.05 - 1	50.05	179	-	0.0500	245.1	PPB	29 Jun 2020 15:21:15	S	Std
0.2 - 1	50.2	683	-	0.2000	245.1	PPB	29 Jun 2020 15:23:32	S	Std
2.5 - 1	52.5	9434	-	2.5000	245.1	PPB	29 Jun 2020 15:25:49	S	Std
5.0 - 1	55	19138	-	5.0000	245.1	PPB	29 Jun 2020 15:28:06	S	Std
7.5 - 1	57.5	27695	-	7.5000	245.1	PPB	29 Jun 2020 15:30:22	S	Std
10.0 - 1	510	37892	-	10.0000	245.1	PPB	29 Jun 2020 15:37:52	S	Std
ICV05 - 1	ICV05	14907	3.9611	-245.1		PPB	29 Jun 2020 15:43:07	U	SMPL
ICB05 - 1	ICB05	-126	-0.0289	-245.1		PPB	29 Jun 2020 15:45:22	U	SMPL
CCV01 - 1	CCV01	18847	5.0068	-245.1		PPB	29 Jun 2020 15:47:37	U	SMPL
CCB01 - 1	CCB01	-127	-0.0292	-245.1		PPB	29 Jun 2020 15:49:52	U	SMPL
CRA - 1	CRA	160	0.0470	-245.1		PPB	29 Jun 2020 15:52:07	U	SMPL
HighStd - 1	HighStd	37544	9.9692	-245.1		PPB	29 Jun 2020 15:54:22	U	SMPL
ChkStd - 1	ChkStd	26070	6.9239	-245.1		PPB	29 Jun 2020 15:56:38	U	SMPL
PB129837BL - 1	PBW	-152	-0.0358	-245.1		PPB	29 Jun 2020 15:58:55	U	SMPL
PB129837BS - 1	LCSW	15587	4.1415	-245.1		PPB	29 Jun 2020 16:01:12	U	SMPL
L3125-01 - 1	TW-2	-135	-0.0313	-245.1		PPB	29 Jun 2020 16:03:29	U	SMPL
L3125-01DUP - 1	TW-2DUP	-73	-0.0148	-245.1		PPB	29 Jun 2020 16:05:45	U	SMPL
L3125-01MS - 1	TW-2MS	13971	3.7126	-245.1		PPB	29 Jun 2020 16:08:02	U	SMPL
L3125-01MSD - 1	TW-2MSD	13981	3.7153	-245.1		PPB	29 Jun 2020 16:10:19	U	SMPL
CCV02 - 1	CCV02	18916	5.0251	-245.1		PPB	29 Jun 2020 16:12:35	U	SMPL
CCB02 - 1	CCB02	-122	-0.0278	-245.1		PPB	29 Jun 2020 16:14:50	U	SMPL
L3125-01LX5 - 1		-60	-0.0114	-245.1		PPB	29 Jun 2020 16:17:05	U	SMPL
L3125-01A - 1		14924	3.9656	-245.1		PPB	29 Jun 2020 16:19:21	U	SMPL
CCV03 - 1	CCV03	18971	5.0397	-245.1		PPB	29 Jun 2020 16:21:36	U	SMPL
CCB03 - 1	CCB03	-140	-0.0326	-245.1		PPB	29 Jun 2020 16:23:52	U	SMPL