



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109787

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/29/2020	16:05
2	CAL2	1	Water	NA	NA	20.1	7.00	06/29/2020	16:06
3	CAL3	1	Water	NA	NA	20.1	10.03	06/29/2020	16:07
4	ICV	1	Water	NA	NA	20.1	7.01	06/29/2020	16:09
5	CCV1	1	Water	NA	NA	20.1	2.01	06/29/2020	16:10
6	L3152-01	1	Water	NA	NA	26.1	8.10	06/29/2020	17:25
7	L3152-01DUP	1	Water	NA	NA	27.2	8.11	06/29/2020	17:26
8	CCV2	1	Water	NA	NA	20.2	12.01	06/29/2020	17:29

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3152

WorkList ID : 140019

Department : Wet-Chemistry

Date : 06-29-2020 16:03:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2020	Water	L3152-01-5	pH	Cool 4 deg C	PSEG01	R41	40680	06/29/2020	9040C

Date/Time

Received by:

Relinquished by:

06/29/2020
for water Chain
CSM

Date/Time

Received by:

Relinquished by:

06/29/2020
for CSM
for Wet Chain