



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109788

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/29/2020	16:05
2	CAL2	1	Water	NA	NA	20.1	7.01	06/29/2020	16:06
3	CAL3	1	Water	NA	NA	20.1	10.03	06/29/2020	16:07
4	ICV	1	Water	NA	NA	20.1	7.00	06/29/2020	16:09
5	CCV1	1	Water	NA	NA	20.1	2.02	06/29/2020	16:10
6	L3152-04	1	Solid	20.03	20	23.3	6.62	06/29/2020	16:11
7	L3152-04DUP	1	Solid	20.04	20	23.4	6.63	06/29/2020	16:12
8	L3152-05	1	Solid	20.02	20	23.0	6.48	06/29/2020	16:15
9	L3152-06	1	Solid	20.03	20	22.6	6.08	06/29/2020	16:16
10	CCV2	1	Water	NA	NA	20.1	12.02	06/29/2020	16:20

WORKLIST(Hardcopy Internal Chain)

✓ 109788

WorkList Name : PH S L3152 WorkList ID : 140020 Department : Wet-Chemistry Date : 06-29-2020 16:03:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2020	Solid	L3152-04	pH	Cool 4 deg C	PSEG01	R41	40855	06/29/2020	9045D
06/29/2020	Solid	L3152-05	pH	Cool 4 deg C	PSEG01	R41	40856	06/29/2020	9045D
06/29/2020	Solid	L3152-06	pH	Cool 4 deg C	PSEG01	R41	40857	06/29/2020	9045D

Date/Time 06/29/2020
 Received by: Jp Wt-Chem
 Relinquished by: Jp Wt-Chem

Date/Time 06/29/2020
 Received by: Jp Wt-Chem
 Relinquished by: Jp Wt-Chem