



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109809

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/30/2020	17:20
2	CAL2	1	Water	NA	NA	20.2	7.00	06/30/2020	17:21
3	CAL3	1	Water	NA	NA	20.2	10.03	06/30/2020	17:22
4	ICV	1	Water	NA	NA	20.2	6.98	06/30/2020	17:25
5	CCV1	1	Water	NA	NA	20.2	2.02	06/30/2020	17:26
6	L2811-08	1	Solid	20.03	20	22.6	5.99	07/01/2020	17:27
7	L2811-08DUP	1	Solid	20.04	20	22.6	6.01	06/30/2020	17:28
8	L2814-19	1	Solid	20.01	20	22.4	5.94	06/30/2020	17:29
9	L2815-20	1	Solid	20.03	20	22.7	7.20	06/30/2020	17:30
10	L3171-01	1	Solid	20.04	20	22.0	7.84	06/30/2020	17:32
11	L3171-02	1	Solid	20.03	20	21.9	7.04	06/30/2020	17:35
12	L3171-03	1	Solid	20.02	20	21.9	7.03	06/30/2020	17:36
13	CCV2	1	Water	NA	NA	20.2	12.02	06/30/2020	17:40

109809

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3157 WorkList ID : 140046 Department : Wet-Chemistry Date : 06-30-2020 11:51:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/30/2020	Solid	L2811-08	Corrosivity	Cool 4 deg C	PSEG05		NB-7-063020	06/30/2020	9045D
06/30/2020	Solid	L2814-19	Corrosivity	Cool 4 deg C	PSEG05		PL-02-063020	06/30/2020	9045D
06/29/2020	Solid	L2815-20	Corrosivity	Cool 4 deg C	PSEG05		OR-03-062920	06/29/2020	9045D
06/29/2020	Solid	L3171-01	Corrosivity	Cool 4 deg C	CLEA11		CHASSIS-WASH	06/29/2020	9045D
06/29/2020	Solid	L3171-02	Corrosivity	Cool 4 deg C	CLEA11		BUS-WASH	06/29/2020	9045D
06/29/2020	Solid	L3171-03	Corrosivity	Cool 4 deg C	CLEA11		RECLAIM-SYSTEM	06/29/2020	9045D

16:40

Date/Time 06/30/2020 16:40

Received by: JA

Relinquished by: RJ EXT-LAB

Date/Time 06/30/2020 18:30

Received by: RJ EXT-LAB

Relinquished by: J