



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109810

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/01/2020	08:25
2	CAL2	1	Water	NA	NA	20.1	7.00	07/01/2020	08:26
3	CAL3	1	Water	NA	NA	20.1	10.03	07/01/2020	08:27
4	ICV	1	Water	NA	NA	20.2	7.00	07/01/2020	08:30
5	CCV1	1	Water	NA	NA	20.1	2.01	07/01/2020	08:32
6	L3162-02	1	Solid	20.03	20	21.7	8.79	07/01/2020	08:35
7	L3162-02DUP	1	Solid	20.04	20	21.6	8.80	07/01/2020	08:36
8	CCV2	1	Water	NA	NA	20.1	12.02	07/01/2020	08:39

LB

LB 109810

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3162

WorkList ID : 140059

Department : Wet-Chemistry

Date : 07-01-2020 08:11:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/30/2020	Solid	L3162-02 -V	pH	Cool 4 deg C	NEWP02	R42	GL-SP	06/30/2020	9045D

Date/Time 07/01/2020 8:15

Received by: Jy

Relinquished by: RJ EXT. Lab

Date/Time 07/01/2020 9:25

Received by: RJ EXT. Lab

Relinquished by: Jy